

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
 Data File : VY008125.D
 Acq On : 30 Mar 2022 21:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 05:04:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	64	0.00
2 T	Dichlorodifluoromethane	0.451	0.389	13.7	51	0.00
3 P	Chloromethane	0.549	0.503	8.4	55	0.00
4 C	Vinyl Chloride	0.730	0.654	10.4#	53	0.00
5 T	Bromomethane	0.598	0.557	6.9	59	0.00
6 T	Chloroethane	0.480	0.449	6.5	56	0.00
7 T	Trichlorofluoromethane	0.924	0.799	13.5	53	0.00
8 T	Diethyl Ether	0.299	0.295	1.3	59	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.518	9.6	54	0.00
10 T	Methyl Iodide	0.592	0.528	10.8	47#	0.01
11 T	Tert butyl alcohol	0.052	0.051	1.9	57	0.00
12 CM	1,1-Dichloroethene	0.530	0.478	9.8#	55	0.00
13 T	Acrolein	0.042	0.047	-11.9	72	0.00
14 T	Allyl chloride	0.698	0.663	5.0	56	0.00
15 T	Acrylonitrile	0.142	0.152	-7.0	62	0.01
16 T	Acetone	0.115	0.104	9.6	51	0.00
17 T	Carbon Disulfide	1.503	1.333	11.3	52	0.00
18 T	Methyl Acetate	0.386	0.320	17.1	62	0.00
19 T	Methyl tert-butyl Ether	1.508	1.490	1.2	59	0.01
20 T	Methylene Chloride	0.673	0.584	13.2	60	0.00
21 T	trans-1,2-Dichloroethene	0.614	0.589	4.1	58	0.00
22 T	Diisopropyl ether	1.502	1.519	-1.1	59	0.00
23 T	Vinyl Acetate	0.930	1.026	-10.3	60	0.00
24 P	1,1-Dichloroethane	0.994	0.980	1.4	59	0.00
25 T	2-Butanone	0.174	0.177	-1.7	58	0.00
26 T	2,2-Dichloropropane	0.941	0.875	7.0	55	0.00
27 T	cis-1,2-Dichloroethene	0.712	0.699	1.8	60	0.00
28 T	Bromochloromethane	0.322	0.263	18.3	60	0.00
29 T	Tetrahydrofuran	0.113	0.119	-5.3	60	0.00
30 C	Chloroform	1.076	1.083	-0.7#	60	0.00
31 T	Cyclohexane	0.938	0.803	14.4	54	0.00
32 T	1,1,1-Trichloroethane	0.995	0.958	3.7	57	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.496	1.2	61	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	65	0.00
35 S	Dibromofluoromethane	0.313	0.304	2.9	62	0.00
36 T	1,1-Dichloropropene	0.492	0.450	8.5	56	0.00
37 T	Ethyl Acetate	0.230	0.230	0.0	59	0.00
38 T	Carbon Tetrachloride	0.537	0.494	8.0	56	0.00
39 T	Methylcyclohexane	0.665	0.582	12.5	54	0.00
40 TM	Benzene	1.409	1.336	5.2	59	0.00
41 T	Methacrylonitrile	0.116	0.117	-0.9	67	0.00
42 TM	1,2-Dichloroethane	0.395	0.390	1.3	61	0.00
43 T	Isopropyl Acetate	0.452	0.451	0.2	60	0.00
44 TM	Trichloroethene	0.407	0.379	6.9	58	0.00
45 C	1,2-Dichloropropane	0.332	0.320	3.6#	59	0.00
46 T	Dibromomethane	0.218	0.218	0.0	61	0.00
47 T	Bromodichloromethane	0.499	0.487	2.4	60	0.00
48 T	Methyl methacrylate	0.200	0.191	4.5	57	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
 Data File : VY008125.D
 Acq On : 30 Mar 2022 21:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 05:04:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	60	0.00
50 S	Toluene-d8	1.181	1.118	5.3	59	0.00
51 T	4-Methyl-2-Pentanone	0.228	0.233	-2.2	60	0.00
52 CM	Toluene	0.924	0.891	3.6#	58	0.00
53 T	t-1,3-Dichloropropene	0.512	0.502	2.0	59	0.00
54 T	cis-1,3-Dichloropropene	0.579	0.565	2.4	59	0.00
55 T	1,1,2-Trichloroethane	0.301	0.306	-1.7	61	0.00
56 T	Ethyl methacrylate	0.393	0.397	-1.0	58	0.00
57 T	1,3-Dichloropropane	0.493	0.499	-1.2	61	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.151	7.4	62	0.00
59 T	2-Hexanone	0.159	0.163	-2.5	59	0.00
60 T	Dibromochloromethane	0.367	0.363	1.1	59	0.00
61 T	1,2-Dibromoethane	0.300	0.302	-0.7	60	0.00
62 S	4-Bromofluorobenzene	0.417	0.393	5.8	59	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
64 T	Tetrachloroethene	0.384	0.347	9.6	58	0.00
65 PM	Chlorobenzene	1.127	1.084	3.8	59	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.407	2.2	59	0.00
67 C	Ethyl Benzene	1.976	1.885	4.6#	58	0.00
68 T	m/p-Xylenes	0.786	0.763	2.9	58	0.00
69 T	o-Xylene	0.752	0.726	3.5	58	0.00
70 T	Styrene	1.232	1.226	0.5	59	0.00
71 P	Bromoform	0.251	0.250	0.4	58	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
73 T	Isopropylbenzene	4.127	3.875	6.1	57	0.00
74 T	N-amyl acetate	0.907	0.876	3.4	59	0.00
75 P	1,1,2,2-Tetrachloroethane	0.802	0.822	-2.5	61	0.00
76 T	1,2,3-Trichloropropane	0.610	0.630	-3.3	66	0.00
77 T	Bromobenzene	0.947	0.899	5.1	57	0.00
78 T	n-propylbenzene	4.868	4.586	5.8	58	0.00
79 T	2-Chlorotoluene	2.700	2.556	5.3	58	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.219	4.8	58	0.00
81 T	trans-1,4-Dichloro-2-butene	0.273	0.271	0.7	58	0.00
82 T	4-Chlorotoluene	2.771	2.619	5.5	58	0.00
83 T	tert-Butylbenzene	3.012	2.857	5.1	57	0.00
84 T	1,2,4-Trimethylbenzene	3.292	3.158	4.1	59	0.00
85 T	sec-Butylbenzene	4.446	4.219	5.1	58	0.00
86 T	p-Isopropyltoluene	3.700	3.529	4.6	57	0.00
87 T	1,3-Dichlorobenzene	1.877	1.839	2.0	58	0.00
88 T	1,4-Dichlorobenzene	1.872	1.796	4.1	58	0.00
89 T	n-Butylbenzene	3.404	3.165	7.0	57	0.00
90 T	Hexachloroethane	0.717	0.681	5.0	57	0.00
91 T	1,2-Dichlorobenzene	1.653	1.622	1.9	58	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.122	4.7	58	0.00
93 T	1,2,4-Trichlorobenzene	0.990	0.889	10.2	55	0.00
94 T	Hexachlorobutadiene	0.524	0.464	11.5	57	0.00
95 T	Naphthalene	2.188	2.067	5.5	57	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
Data File : VY008125.D
Acq On : 30 Mar 2022 21:09
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 31 05:04:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.856	0.776	9.3	56	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6