

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033125\
 Data File : VY021714.D
 Acq On : 31 Mar 2025 13:04
 Operator : SY/MD
 Sample : Q1657-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11998RE

Quant Time: Mar 31 15:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	116458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	179259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	138174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	43525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	48847	38.713	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.420%
35) Dibromofluoromethane	7.634	113	58000	49.879	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.760%
50) Toluene-d8	10.109	98	208669	47.170	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.340%
62) 4-Bromofluorobenzene	12.408	95	47882	32.000	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	64.000%
Target Compounds						
16) Acetone	3.860	43	1184	4.585	ug/l	Qvalue # 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033125\
Data File : VY021714.D
Acq On : 31 Mar 2025 13:04
Operator : SY/MD
Sample : Q1657-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11998RE

Quant Time: Mar 31 15:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

