

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033125\
 Data File : VY021715.D
 Acq On : 31 Mar 2025 13:27
 Operator : SY/MD
 Sample : Q1679-03
 Misc : 6.20g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-10-VOC

Quant Time: Apr 01 05:36:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	225472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	427834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	403099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	165889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	148563	60.814	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.620%	
35) Dibromofluoromethane	7.634	113	149548	53.886	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.780%	
50) Toluene-d8	10.109	98	528244	50.032	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.060%	
62) 4-Bromofluorobenzene	12.408	95	168484	47.178	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.360%	
Target Compounds						
16) Acetone	3.866	43	6189	12.379	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033125\
Data File : VY021715.D
Acq On : 31 Mar 2025 13:27
Operator : SY/MD
Sample : Q1679-03
Misc : 6.20g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-10-VOC

Quant Time: Apr 01 05:36:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

