



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033125\
Data File : VY021709.D
Acq On : 31 Mar 2025 09:37
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Title : SW846 8260
Last Update : Fri Mar 28 02:30:29 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	19320	PASS
75	95	30	60	53.8	50832	PASS
95	95	100	100	100.0	94491	PASS
96	95	5	9	6.8	6462	PASS
173	174	0.00	2	1.7	1358	PASS
174	95	50	100	82.9	78312	PASS
175	174	5	9	7.9	6153	PASS
176	174	95	101	96.6	75685	PASS
177	176	5	9	6.7	5035	PASS