

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040122\
 Data File : VY008161.D
 Acq On : 01 Apr 2022 15:17
 Operator : SY/MD
 Sample : N2220-04
 Misc : 4.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A11105

Quant Time: Apr 02 03:57:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	19425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	35695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	21287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	17275	88.639	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 177.280%#	
35) Dibromofluoromethane	7.716	113	14531	64.929	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 129.860%	
50) Toluene-d8	10.179	98	39283	46.590	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.180%	
62) 4-Bromofluorobenzene	12.483	95	5813	19.541	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 39.080%	
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	2521	7.087	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040122\
Data File : VY008161.D
Acq On : 01 Apr 2022 15:17
Operator : SY/MD
Sample : N2220-04
Misc : 4.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A11105

Quant Time: Apr 02 03:57:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

