

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040125\
 Data File : VY021737.D
 Acq On : 01 Apr 2025 13:22
 Operator : SY/MD
 Sample : Q1675-09
 Misc : 8.15g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T3

Quant Time: Apr 02 03:23:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	259184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	488873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	462972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	201176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	169059	60.202	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.400%
35) Dibromofluoromethane	7.634	113	168115	53.012	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.020%
50) Toluene-d8	10.103	98	605610	50.198	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.400%
62) 4-Bromofluorobenzene	12.408	95	206234	50.539	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.080%
Target Compounds						
16) Acetone	3.867	43	9363	16.291	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040125\
Data File : VY021737.D
Acq On : 01 Apr 2025 13:22
Operator : SY/MD
Sample : Q1675-09
Misc : 8.15g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T3

Quant Time: Apr 02 03:23:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

