

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040224\
 Data File : VY017813.D
 Acq On : 02 Apr 2024 09:27
 Operator : SY/MD
 Sample : VY0402SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0402SBL01

Quant Time: Apr 03 02:01:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	185402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	341590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	270473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	88212	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	82157	51.209	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.420%
35) Dibromofluoromethane	7.722	113	103383	52.376	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.760%
50) Toluene-d8	10.191	98	393805	53.410	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.820%
62) 4-Bromofluorobenzene	12.489	95	88990	39.669	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040224\
Data File : VY017813.D
Acq On : 02 Apr 2024 09:27
Operator : SY/MD
Sample : VY0402SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0402SBL01

Quant Time: Apr 03 02:01:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

