


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040225\
Data File : VY021747.D
Acq On : 02 Apr 2025 09:12
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Title : SW846 8260
Last Update : Fri Mar 28 02:30:29 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	20677	PASS
75	95	30	60	53.1	53672	PASS
95	95	100	100	100.0	101037	PASS
96	95	5	9	7.0	7093	PASS
173	174	0.00	2	1.7	1414	PASS
174	95	50	100	81.1	81923	PASS
175	174	5	9	8.3	6788	PASS
176	174	95	101	95.7	78435	PASS
177	176	5	9	6.7	5245	PASS