

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040422\
 Data File : VY008191.D
 Acq On : 04 Apr 2022 19:23
 Operator : SY/MD
 Sample : N2237-02
 Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A13015

Quant Time: Apr 05 04:43:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	55757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	91361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	52521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	10879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	31730	56.720	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.440%	
35) Dibromofluoromethane	7.722	113	32641	56.984	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.960%	
50) Toluene-d8	10.178	98	105604	48.935	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 97.860%	
62) 4-Bromofluorobenzene	12.483	95	15902	20.886	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 41.780%	
Target Compounds						
					Qvalue	
30) Chloroform	7.508	83	2062	1.718	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040422\
Data File : VY008191.D
Acq On : 04 Apr 2022 19:23
Operator : SY/MD
Sample : N2237-02
Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A13015

Quant Time: Apr 05 04:43:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

