

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040425\
 Data File : VY021781.D
 Acq On : 04 Apr 2025 13:27
 Operator : SY/MD
 Sample : Q1730-01
 Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-VSL-15-040325

Quant Time: Apr 04 23:39:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	343368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	647395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	573735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	228826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	179805	48.331	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.660%
35) Dibromofluoromethane	7.634	113	200979	47.857	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.720%
50) Toluene-d8	10.109	98	782529	48.981	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.960%
62) 4-Bromofluorobenzene	12.408	95	240375	44.481	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.960%
Target Compounds						
16) Acetone	3.872	43	12606	16.556	ug/l	99
37) Ethyl Acetate	6.988	43	13216	4.378	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040425\
Data File : VY021781.D
Acq On : 04 Apr 2025 13:27
Operator : SY/MD
Sample : Q1730-01
Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-VSL-15-040325

Quant Time: Apr 04 23:39:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

