

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040425\
 Data File : VY021786.D
 Acq On : 04 Apr 2025 15:24
 Operator : SY/MD
 Sample : Q1730-11
 Misc : 5.90g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-23-040325

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 04/07/2025
 Supervised By :Semsettin Yesilyurt 04/07/2025

Quant Time: Apr 04 23:40:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	341604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	662745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	606680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	250886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	200040	54.048	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 108.100%			
35) Dibromofluoromethane	7.634	113	213518	49.666	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 99.340%			
50) Toluene-d8	10.109	98	813418	49.735	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 99.460%			
62) 4-Bromofluorobenzene	12.408	95	260762	47.136	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 94.280%			
Target Compounds						Qvalue
16) Acetone	3.867	43	22926	30.266	ug/l	95
37) Ethyl Acetate	6.994	43	7123m	2.305	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040425\
Data File : VY021786.D
Acq On : 04 Apr 2025 15:24
Operator : SY/MD
Sample : Q1730-11
Misc : 5.90g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-23-040325

Quant Time: Apr 04 23:40:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 04/07/2025
Supervised By :Semsettin Yesilyurt 04/07/2025

