

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
 Data File : VY008205.D
 Acq On : 05 Apr 2022 13:30
 Operator : SY/MD
 Sample : N2220-17RE
 Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A09145RE

Quant Time: Apr 06 03:56:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	90421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	106111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	31820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2859	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	20918	23.058	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	46.120%#
35) Dibromofluoromethane	7.722	113	36925	55.502	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.000%
50) Toluene-d8	10.179	98	103079	41.125	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.260%
62) 4-Bromofluorobenzene	12.483	95	5827	6.589	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	13.180%#
Target Compounds						Qvalue
32) 1,1,1-Trichloroethane	7.710	97	3516	1.955	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
Data File : VY008205.D
Acq On : 05 Apr 2022 13:30
Operator : SY/MD
Sample : N2220-17RE
Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A09145RE

Quant Time: Apr 06 03:56:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

