

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
 Data File : VY008210.D
 Acq On : 05 Apr 2022 15:27
 Operator : SY/MD
 Sample : N2237-05RE
 Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A12040RE

Quant Time: Apr 06 03:57:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	115289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	203208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	154287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	36565	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	64622	55.868	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.740%
35) Dibromofluoromethane	7.716	113	65895	51.721	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.440%
50) Toluene-d8	10.179	98	243502	50.729	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.483	95	51428	30.369	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
Data File : VY008210.D
Acq On : 05 Apr 2022 15:27
Operator : SY/MD
Sample : N2237-05RE
Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A12040RE

Quant Time: Apr 06 03:57:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

