

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
 Data File : VY008213.D
 Acq On : 05 Apr 2022 16:37
 Operator : SY/MD
 Sample : N2237-10RE
 Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUP-0401RE

Quant Time: Apr 06 03:58:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	107577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	192873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	153264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	41743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	62037	57.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.960%
35) Dibromofluoromethane	7.716	113	61585	50.928	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.860%
50) Toluene-d8	10.179	98	223927	49.151	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.483	95	57158	35.561	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.120%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	38317	28.506	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040522\
Data File : VY008213.D
Acq On : 05 Apr 2022 16:37
Operator : SY/MD
Sample : N2237-10RE
Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUP-0401RE

Quant Time: Apr 06 03:58:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

