

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040524\
 Data File : VY017868.D
 Acq On : 05 Apr 2024 14:07
 Operator : SY/MD
 Sample : P1996-04RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP2RE

Quant Time: Apr 06 04:47:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5124	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	3445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	1913	62.555	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.120%
35) Dibromofluoromethane	7.716	113	1022	34.517	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	69.040%
50) Toluene-d8	10.191	98	5509	49.809	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.620%
62) 4-Bromofluorobenzene	12.495	95	755	22.437	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	44.880%
Target Compounds						
44) Trichloroethene	8.947	130	395	11.403	ug/l	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040524\
Data File : VY017868.D
Acq On : 05 Apr 2024 14:07
Operator : SY/MD
Sample : P1996-04RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP2RE

Quant Time: Apr 06 04:47:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

