

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040524\
 Data File : VY017869.D
 Acq On : 05 Apr 2024 14:30
 Operator : SY/MD
 Sample : P1980-03
 Misc : 4.23g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC13-GRAB-14-15

Quant Time: Apr 06 04:47:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	113496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	215039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	189020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	67038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	61283	62.399	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.800%
35) Dibromofluoromethane	7.721	113	70296	56.572	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.140%
50) Toluene-d8	10.184	98	253427	54.599	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.200%
62) 4-Bromofluorobenzene	12.489	95	61418	43.491	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.980%
Target Compounds						
20) Methylene Chloride	4.716	84	15785	10.596	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040524\
Data File : VY017869.D
Acq On : 05 Apr 2024 14:30
Operator : SY/MD
Sample : P1980-03
Misc : 4.23g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC13-GRAB-14-15

Quant Time: Apr 06 04:47:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

