

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040622\
 Data File : VY008234.D
 Acq On : 06 Apr 2022 14:57
 Operator : SY/MD
 Sample : N2275-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-03-040522

Quant Time: Apr 07 02:12:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	111714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	206062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	199966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	64267	57.339	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.680%
35) Dibromofluoromethane	7.715	113	65099	50.388	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.780%
50) Toluene-d8	10.178	98	251877	51.747	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.483	95	88008	51.249	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.500%
Target Compounds						
					Qvalue	
69) o-Xylene	12.026	106	3153	1.048	ug/l	91
80) 1,3,5-Trimethylbenzene	12.812	105	39515	5.979	ug/l	99
84) 1,2,4-Trimethylbenzene	13.123	105	16806	2.611	ug/l	97
86) p-Isopropyltoluene	13.367	119	10930	1.511	ug/l	96
89) n-Butylbenzene	13.696	91	11721	1.761	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040622\
Data File : VY008234.D
Acq On : 06 Apr 2022 14:57
Operator : SY/MD
Sample : N2275-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-03-040522

Quant Time: Apr 07 02:12:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

