

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040622\
 Data File : VY008236.D
 Acq On : 06 Apr 2022 15:44
 Operator : SY/MD
 Sample : N2281-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-040622

Quant Time: Apr 07 02:12:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	44955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	78714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	71157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	29827	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	24665	54.686	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.380%
35) Dibromofluoromethane	7.716	113	24584	49.814	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.620%
50) Toluene-d8	10.179	98	93258	50.157	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.320%
62) 4-Bromofluorobenzene	12.483	95	29021	44.241	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040622\
Data File : VY008236.D
Acq On : 06 Apr 2022 15:44
Operator : SY/MD
Sample : N2281-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-040622

Quant Time: Apr 07 02:12:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

