

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040722\
 Data File : VY008248.D
 Acq On : 07 Apr 2022 15:47
 Operator : KP/MD
 Sample : N2272-05
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Apr 08 02:35:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	22580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	38845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	35816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14623	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	13409	59.189	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.380%
35) Dibromofluoromethane	7.716	113	12957	53.201	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.400%
50) Toluene-d8	10.179	98	46342	50.505	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.020%
62) 4-Bromofluorobenzene	12.483	95	14471	44.702	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040722\
Data File : VY008248.D
Acq On : 07 Apr 2022 15:47
Operator : KP/MD
Sample : N2272-05
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Apr 08 02:35:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

