

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040725\
 Data File : VY021794.D
 Acq On : 07 Apr 2025 09:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 03:30:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.518	0.452	12.7	84	0.00
3 P	Chloromethane	0.729	0.643	11.8	88	0.00
4 C	Vinyl Chloride	0.824	0.748	9.2#	88	0.00
5 T	Bromomethane	0.562	0.461	18.0	85	0.00
6 T	Chloroethane	0.539	0.488	9.5	90	0.00
7 T	Trichlorofluoromethane	1.056	0.976	7.6	91	0.00
8 T	Diethyl Ether	0.287	0.250	12.9	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.516	7.4	91	0.00
10 T	Methyl Iodide	0.611	0.600	1.8	86	0.00
11 T	Tert butyl alcohol	0.035	0.028	20.0	77	0.00
12 CM	1,1-Dichloroethene	0.524	0.492	6.1#	90	0.00
13 T	Acrolein	0.064	0.019	70.3#	29#	0.00
14 T	Allyl chloride	0.836	0.773	7.5	87	0.00
15 T	Acrylonitrile	0.120	0.104	13.3	81	0.00
16 T	Acetone	0.111	0.115	-3.6	101	0.00
17 T	Carbon Disulfide	1.727	1.557	9.8	87	0.00
18 T	Methyl Acetate	0.275	0.273	0.7	97	0.00
19 T	Methyl tert-butyl Ether	1.306	1.179	9.7	84	0.00
20 T	Methylene Chloride	0.626	0.524	16.3	88	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.548	5.0	91	0.00
22 T	Diisopropyl ether	1.754	1.671	4.7	88	0.00
23 T	Vinyl Acetate	1.026	0.906	11.7	80	0.00
24 P	1,1-Dichloroethane	1.049	0.993	5.3	90	0.00
25 T	2-Butanone	0.158	0.145	8.2	86	0.00
26 T	2,2-Dichloropropane	0.933	0.900	3.5	92	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.623	4.0	90	0.00
28 T	Bromochloromethane	0.442	0.412	6.8	91	0.00
29 T	Tetrahydrofuran	0.102	0.088	13.7	79	0.00
30 C	Chloroform	1.091	1.038	4.9#	91	0.00
31 T	Cyclohexane	0.956	0.852	10.9	88	0.00
32 T	1,1,1-Trichloroethane	0.986	0.931	5.6	91	0.00
33 S	1,2-Dichloroethane-d4	0.542	0.510	5.9	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.324	0.328	-1.2	88	0.00
36 T	1,1-Dichloropropene	0.494	0.499	-1.0	91	0.00
37 T	Ethyl Acetate	0.233	0.206	11.6	80	0.00
38 T	Carbon Tetrachloride	0.566	0.574	-1.4	93	0.00
39 T	Methylcyclohexane	0.590	0.577	2.2	85	0.00
40 TM	Benzene	1.479	1.489	-0.7	91	0.00
41 T	Methacrylonitrile	0.124	0.122	1.6	81	-0.01
42 TM	1,2-Dichloroethane	0.417	0.403	3.4	87	0.00
43 T	Isopropyl Acetate	0.442	0.397	10.2	80	0.00
44 TM	Trichloroethene	0.377	0.391	-3.7	95	0.00
45 C	1,2-Dichloropropane	0.350	0.349	0.3#	90	0.00
46 T	Dibromomethane	0.202	0.194	4.0	89	0.00
47 T	Bromodichloromethane	0.523	0.517	1.1	90	0.00
48 T	Methyl methacrylate	0.203	0.192	5.4	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040725\
 Data File : VY021794.D
 Acq On : 07 Apr 2025 09:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 03:30:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	83	0.00
50 S	Toluene-d8	1.234	1.277	-3.5	89	0.00
51 T	4-Methyl-2-Pentanone	0.230	0.214	7.0	81	0.00
52 CM	Toluene	0.923	0.947	-2.6#	91	0.00
53 T	t-1,3-Dichloropropene	0.465	0.463	0.4	87	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.550	-0.4	89	0.00
55 T	1,1,2-Trichloroethane	0.260	0.255	1.9	89	0.00
56 T	Ethyl methacrylate	0.332	0.324	2.4	81	0.00
57 T	1,3-Dichloropropane	0.449	0.442	1.6	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.162	0.167	-3.1	85	0.00
59 T	2-Hexanone	0.153	0.149	2.6	83	0.00
60 T	Dibromochloromethane	0.355	0.353	0.6	88	0.00
61 T	1,2-Dibromoethane	0.244	0.237	2.9	85	0.00
62 S	4-Bromofluorobenzene	0.417	0.433	-3.8	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.455	0.498	-9.5	101	0.00
65 PM	Chlorobenzene	1.146	1.144	0.2	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.406	0.2	89	0.00
67 C	Ethyl Benzene	1.965	2.032	-3.4#	90	0.00
68 T	m/p-Xylenes	0.753	0.782	-3.9	90	0.00
69 T	o-Xylene	0.695	0.722	-3.9	89	0.00
70 T	Styrene	1.158	1.220	-5.4	90	0.00
71 P	Bromoform	0.232	0.230	0.9	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.714	3.754	-1.1	90	0.00
74 T	N-amyl acetate	0.872	0.752	13.8	75	0.00
75 P	1,1,2,2-Tetrachloroethane	0.668	0.558	16.5	78	0.00
76 T	1,2,3-Trichloropropane	0.496	0.495	0.2	91	0.00
77 T	Bromobenzene	0.892	0.880	1.3	90	0.00
78 T	n-propylbenzene	4.465	4.571	-2.4	90	0.00
79 T	2-Chlorotoluene	2.593	2.571	0.8	89	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.067	-1.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.199	7.4	82	0.00
82 T	4-Chlorotoluene	2.704	2.643	2.3	87	0.00
83 T	tert-Butylbenzene	2.703	2.671	1.2	87	0.00
84 T	1,2,4-Trimethylbenzene	3.000	3.088	-2.9	90	0.00
85 T	sec-Butylbenzene	3.951	3.994	-1.1	89	0.00
86 T	p-Isopropyltoluene	3.265	3.370	-3.2	90	0.00
87 T	1,3-Dichlorobenzene	1.757	1.749	0.5	91	0.00
88 T	1,4-Dichlorobenzene	1.736	1.688	2.8	89	0.00
89 T	n-Butylbenzene	3.020	3.049	-1.0	88	0.00
90 T	Hexachloroethane	0.716	0.693	3.2	89	0.00
91 T	1,2-Dichlorobenzene	1.529	1.506	1.5	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.093	9.7	83	0.00
93 T	1,2,4-Trichlorobenzene	0.829	0.817	1.4	85	0.00
94 T	Hexachlorobutadiene	0.518	0.498	3.9	87	0.00
95 T	Naphthalene	1.375	1.295	5.8	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040725\
Data File : VY021794.D
Acq On : 07 Apr 2025 09:10
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 03:30:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.707	0.687	2.8	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6