

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041025\
 Data File : VY021827.D
 Acq On : 10 Apr 2025 10:28
 Operator : SY/MD
 Sample : VY0410SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0410SBL01

Quant Time: Apr 11 01:49:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	241941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	436661	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	362482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	131635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	122060	46.564	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.120%
35) Dibromofluoromethane	7.640	113	141417	49.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.860%
50) Toluene-d8	10.109	98	522401	48.479	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.960%
62) 4-Bromofluorobenzene	12.408	95	139589	38.297	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041025\
Data File : VY021827.D
Acq On : 10 Apr 2025 10:28
Operator : SY/MD
Sample : VY0410SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0410SBL01

Quant Time: Apr 11 01:49:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

