

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041122\
 Data File : VY008276.D
 Acq On : 11 Apr 2022 16:35
 Operator : KP/MD
 Sample : N2340-08
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-08-17.5-18.0

Quant Time: Apr 12 06:59:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	82539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	140450	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	121237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	51838	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	37003	47.727	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.460%
35) Dibromofluoromethane	7.716	113	40817	50.316	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.640%
50) Toluene-d8	10.179	98	165072	59.468	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	118.940%
62) 4-Bromofluorobenzene	12.483	95	50870	48.724	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.440%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041122\
Data File : VY008276.D
Acq On : 11 Apr 2022 16:35
Operator : KP/MD
Sample : N2340-08
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-08-17.5-18.0

Quant Time: Apr 12 06:59:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35




Instrument :
MSVOA_Y
ClientSampleId :
SB-08-17.5-18.0


Tgt Ion:168 Resp: 82539
Ion Ratio Lower Upper
168 100
99 54.6 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 47.727 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. 0.001 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35

Tgt Ion: 65 Resp: 37003
Ion Ratio Lower Upper
65 100
67 53.9 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008276.D

Acq: 11 Apr 2022 16:35

Instrument :

MSVOA_Y

ClientSampleId :

SB-08-17.5-18.0





Tgt Ion:114 Resp: 140450

Ion Ratio Lower Upper

114 100

63 16.3 0.0 43.2

88 15.0 0.0 31.8



#35

Dibromofluoromethane

Concen: 50.316 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.001 min

Lab File: VY008276.D

Acq: 11 Apr 2022 16:35




Tgt Ion:113 Resp: 40817


Ion Ratio Lower Upper

113 100

111 104.1 82.6 124.0

192 18.6 17.6 26.4









#50
Toluene-d8
Concen: 59.468 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.001 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35




Instrument :
MSVOA_Y
ClientSampleId :
SB-08-17.5-18.0


Tgt Ion: 98 Resp: 165072
Ion Ratio Lower Upper
98 100
100 65.1 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 48.724 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35

Tgt Ion: 95 Resp: 50870
Ion Ratio Lower Upper
95 100
174 82.6 0.0 177.0
176 81.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35

Instrument :
MSVOA_Y
ClientSampleId :
SB-08-17.5-18.0

Tgt Ion:117 Resp: 121237
Ion Ratio Lower Upper
117 100
82 51.8 43.0 64.4
119 32.7 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008276.D
Acq: 11 Apr 2022 16:35

Tgt Ion:152 Resp: 51838
Ion Ratio Lower Upper
152 100
115 59.5 28.2 84.7
150 159.5 0.0 347.6

