

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
 Data File : VY013235.D
 Acq On : 10 Apr 2023 18:55
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Apr 11 06:43:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 06:41:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	67887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	113228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	104487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	48304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	47789	62.859	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.720%
35) Dibromofluoromethane	7.716	113	37859	55.694	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.380%
50) Toluene-d8	10.179	98	135860	57.049	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.100%
62) 4-Bromofluorobenzene	12.483	95	42525	50.628	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.260%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	13649	11.621	ug/l	93
93) 1,2,4-Trichlorobenzene	15.007	180	1466	1.741	ug/l	100
94) Hexachlorobutadiene	15.111	225	511	1.024	ug/l	91
95) Naphthalene	15.239	128	3586	1.911	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	1558	1.996	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
Data File : VY013235.D
Acq On : 10 Apr 2023 18:55
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Apr 11 06:43:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 06:41:51 2023
Response via : Initial Calibration

