

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
 Data File : VY013243.D
 Acq On : 11 Apr 2023 15:27
 Operator : KP/MD
 Sample : 02230-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-040623

Quant Time: Apr 11 17:10:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 11:59:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	10641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	15505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	14320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	6933	57.938	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.880%
35) Dibromofluoromethane	7.722	113	5875	63.389	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.780%
50) Toluene-d8	10.179	98	19003	58.563	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.120%
62) 4-Bromofluorobenzene	12.483	95	5097	43.988	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.980%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.698	84	2888	19.690	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
Data File : VY013243.D
Acq On : 11 Apr 2023 15:27
Operator : KP/MD
Sample : 02230-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-040623

Quant Time: Apr 11 17:10:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 11:59:49 2023
Response via : Initial Calibration

