

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
 Data File : VY013244.D
 Acq On : 11 Apr 2023 15:51
 Operator : KP/MD
 Sample : 02208-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-124035

Quant Time: Apr 11 17:10:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 11:59:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	62070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	103194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	92144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	38736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	40995	58.732	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 117.460%			
35) Dibromofluoromethane	7.716	113	34826	56.458	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 112.920%			
50) Toluene-d8	10.179	98	127084	58.845	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 117.680%			
62) 4-Bromofluorobenzene	12.483	95	35395	45.896	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 91.800%			
Target Compounds						
16) Acetone	3.954	43	3810	13.413	ug/l	90
20) Methylene Chloride	4.710	84	11545	10.030	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
Data File : VY013244.D
Acq On : 11 Apr 2023 15:51
Operator : KP/MD
Sample : 02208-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-124035

Quant Time: Apr 11 17:10:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 11:59:49 2023
Response via : Initial Calibration

