

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
 Data File : VY013252.D
 Acq On : 11 Apr 2023 18:57
 Operator : KP/MD
 Sample : 02270-02
 Misc : 4.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-17-7-9-20230410

Quant Time: Apr 12 01:37:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 11:59:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	63367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	109018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	105058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	49229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	45755	64.210	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 128.420%		
35) Dibromofluoromethane	7.716	113	38689	59.370	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 118.740%		
50) Toluene-d8	10.179	98	135290	59.298	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 118.600%		
62) 4-Bromofluorobenzene	12.483	95	44105	54.135	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 108.260%		
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	9803	6.490	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041123\
Data File : VY013252.D
Acq On : 11 Apr 2023 18:57
Operator : KP/MD
Sample : 02270-02
Misc : 4.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-17-7-9-20230410

Quant Time: Apr 12 01:37:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041123S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 11:59:49 2023
Response via : Initial Calibration

