

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041322\
 Data File : VY008310.D
 Acq On : 13 Apr 2022 14:16
 Operator : KP/MD
 Sample : N2364-05RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-14.5-15.0RE

Quant Time: Apr 14 05:56:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	52298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	91894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	84389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	35885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	28168	57.340	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	114.680%	
35) Dibromofluoromethane	7.716	113	28426	53.557	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	107.120%	
50) Toluene-d8	10.179	98	110122	60.634	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	121.260%	
62) 4-Bromofluorobenzene	12.483	95	34656	50.733	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	101.460%	
Target Compounds						
16) Acetone	3.948	43	1542	13.050	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041322\
Data File : VY008310.D
Acq On : 13 Apr 2022 14:16
Operator : KP/MD
Sample : N2364-05RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-14.5-15.0RE

Quant Time: Apr 14 05:56:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

