

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041322\
 Data File : VY008311.D
 Acq On : 13 Apr 2022 14:40
 Operator : KP/MD
 Sample : N2364-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-14.5-15.0RE

Quant Time: Apr 14 05:56:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	914	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	1701	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	726	84.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 169.120%#	
35) Dibromofluoromethane	7.716	113	420	42.749	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 85.500%	
50) Toluene-d8	10.185	98	1843	54.822	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.640%	
62) 4-Bromofluorobenzene	12.489	95	333	26.335	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 52.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041322\
Data File : VY008311.D
Acq On : 13 Apr 2022 14:40
Operator : KP/MD
Sample : N2364-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-14.5-15.0RE

Quant Time: Apr 14 05:56:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

