

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041422\
 Data File : VY008328.D
 Acq On : 14 Apr 2022 15:26
 Operator : KP/MD
 Sample : N2349-15RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DEL-8RE

Quant Time: Apr 15 06:14:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	177739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	145459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	45659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53112	54.169	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.340%	
35) Dibromofluoromethane	7.716	113	52909	51.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.080%	
50) Toluene-d8	10.179	98	205503	58.501	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 117.000%	
62) 4-Bromofluorobenzene	12.483	95	54610	41.332	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 82.660%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	10500	44.524	ug/l #	84
25) 2-Butanone	6.990	43	5950	16.135	ug/l #	80
51) 4-Methyl-2-Pentanone	10.069	43	6241	7.704	ug/l	90
59) 2-Hexanone	10.831	43	5799	10.404	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041422\
Data File : VY008328.D
Acq On : 14 Apr 2022 15:26
Operator : KP/MD
Sample : N2349-15RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DEL-8RE

Quant Time: Apr 15 06:14:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

