

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041422\
 Data File : VY008333.D
 Acq On : 14 Apr 2022 17:25
 Operator : KP/MD
 Sample : N2401-03
 Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-3-11.5

Quant Time: Apr 15 06:16:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	106358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	193847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	184972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	82969	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	61887	61.946	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.900%
35) Dibromofluoromethane	7.716	113	60799	54.303	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.600%
50) Toluene-d8	10.179	98	231790	60.502	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	121.000%
62) 4-Bromofluorobenzene	12.483	95	81549	56.593	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041422\
Data File : VY008333.D
Acq On : 14 Apr 2022 17:25
Operator : KP/MD
Sample : N2401-03
Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-3-11.5

Quant Time: Apr 15 06:16:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

