

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041425\
 Data File : VY021873.D
 Acq On : 14 Apr 2025 16:37
 Operator : SY/MD
 Sample : Q1791-07
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 290502

Quant Time: Apr 15 02:47:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	339985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	608881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	512429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	209894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	175549	47.656	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.320%
35) Dibromofluoromethane	7.634	113	187340	47.431	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.860%
50) Toluene-d8	10.109	98	709480	47.217	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.440%
62) 4-Bromofluorobenzene	12.407	95	220234	43.332	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.660%
Target Compounds						
						Qvalue
16) Acetone	3.879	43	24799	32.894	ug/l	93
25) 2-Butanone	6.915	43	1429	1.327	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041425\
 Data File : VY021873.D
 Acq On : 14 Apr 2025 16:37
 Operator : SY/MD
 Sample : Q1791-07
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 290502

Quant Time: Apr 15 02:47:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

