

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041525\
 Data File : VY021894.D
 Acq On : 15 Apr 2025 14:26
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-15-VOC

Quant Time: Apr 16 03:02:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	253946	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	464077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	390210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	144267	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	126571	46.002	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.000%
35) Dibromofluoromethane	7.634	113	148668	49.385	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.760%
50) Toluene-d8	10.109	98	553750	48.352	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.408	95	152165	39.281	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041525\
Data File : VY021894.D
Acq On : 15 Apr 2025 14:26
Operator : SY/MD
Sample : IBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-15-VOC

Quant Time: Apr 16 03:02:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY021894.D
Acq: 15 Apr 2025 14:26




Instrument :
MSVOA_Y
ClientSampleId :
TP-15-VOC


Tgt Ion:168 Resp: 253946
Ion Ratio Lower Upper
168 100
99 56.2 46.7 70.1



#33
1,2-Dichloroethane-d4
Concen: 46.002 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY021894.D
Acq: 15 Apr 2025 14:26

Tgt Ion: 65 Resp: 126571
Ion Ratio Lower Upper
65 100
67 52.4 0.0 104.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VY021894.D
Acq: 15 Apr 2025 14:26




Instrument :
MSVOA_Y
ClientSampleId :
TP-15-VOC


Tgt Ion:114 Resp: 464077
Ion Ratio Lower Upper
114 100
63 20.6 0.0 40.2
88 15.0 0.0 29.8



#35
Dibromofluoromethane
Concen: 49.385 ug/l
RT: 7.634 min Scan# 970
Delta R.T. 0.000 min
Lab File: VY021894.D
Acq: 15 Apr 2025 14:26

Tgt Ion:113 Resp: 148668
Ion Ratio Lower Upper
113 100
111 103.1 82.6 123.8
192 19.9 16.2 24.4





#50

Toluene-d8

Concen: 48.352 ug/l

RT: 10.109 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY021894.D

Acq: 15 Apr 2025 14:26

Instrument :

MSVOA_Y

ClientSampleId :





TP-15-VOC

Tgt Ion: 98 Resp: 553750

Ion Ratio Lower Upper

98 100

100 64.5 52.1 78.1



#62

4-Bromofluorobenzene

Concen: 39.281 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. 0.000 min

Lab File: VY021894.D

Acq: 15 Apr 2025 14:26




Tgt Ion: 95 Resp: 152165


Ion Ratio Lower Upper

95 100

174 82.9 0.0 170.8

176 80.3 0.0 162.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY021894.D

Acq: 15 Apr 2025 14:26

Instrument :

MSVOA_Y

ClientSampleId :

TP-15-VOC





Tgt Ion:117 Resp: 390210

Ion Ratio Lower Upper

117 100

82 55.2 42.8 64.2

119 31.5 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY021894.D

Acq: 15 Apr 2025 14:26

Tgt Ion:152 Resp: 144267

Ion Ratio Lower Upper

152 100

115 58.1 28.8 86.5

150 156.9 0.0 348.4

