

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041725\
 Data File : VY021900.D
 Acq On : 17 Apr 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 18 03:52:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.518	0.448	13.5	82	0.00
3 P	Chloromethane	0.729	0.610	16.3	82	0.00
4 C	Vinyl Chloride	0.824	0.708	14.1#	82	0.00
5 T	Bromomethane	0.562	0.479	14.8	87	0.00
6 T	Chloroethane	0.539	0.479	11.1	87	0.00
7 T	Trichlorofluoromethane	1.056	0.986	6.6	90	0.00
8 T	Diethyl Ether	0.287	0.260	9.4	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.542	2.7	94	0.00
10 T	Methyl Iodide	0.611	0.631	-3.3	89	0.00
11 T	Tert butyl alcohol	0.035	0.029	17.1	81	0.02
12 CM	1,1-Dichloroethene	0.524	0.491	6.3#	89	0.00
13 T	Acrolein	0.064	0.011	82.8#	16#	0.01
14 T	Allyl chloride	0.836	0.767	8.3	85	0.00
15 T	Acrylonitrile	0.120	0.111	7.5	85	0.00
16 T	Acetone	0.111	0.099	10.8	85	0.00
17 T	Carbon Disulfide	1.727	1.504	12.9	82	0.00
18 T	Methyl Acetate	0.275	0.326	-18.5	114	0.00
19 T	Methyl tert-butyl Ether	1.306	1.243	4.8	87	0.00
20 T	Methylene Chloride	0.626	0.544	13.1	89	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.553	4.2	90	0.00
22 T	Diisopropyl ether	1.754	1.709	2.6	88	0.00
23 T	Vinyl Acetate	1.026	0.933	9.1	81	0.00
24 P	1,1-Dichloroethane	1.049	1.005	4.2	90	0.00
25 T	2-Butanone	0.158	0.144	8.9	84	0.00
26 T	2,2-Dichloropropane	0.933	0.908	2.7	92	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.639	1.5	91	0.00
28 T	Bromochloromethane	0.442	0.422	4.5	92	0.00
29 T	Tetrahydrofuran	0.102	0.092	9.8	81	0.00
30 C	Chloroform	1.091	1.055	3.3#	91	0.00
31 T	Cyclohexane	0.956	0.862	9.8	87	0.00
32 T	1,1,1-Trichloroethane	0.986	0.962	2.4	92	0.00
33 S	1,2-Dichloroethane-d4	0.542	0.489	9.8	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.324	0.325	-0.3	85	0.00
36 T	1,1-Dichloropropene	0.494	0.501	-1.4	89	0.00
37 T	Ethyl Acetate	0.233	0.220	5.6	83	0.00
38 T	Carbon Tetrachloride	0.566	0.594	-4.9	93	0.00
39 T	Methylcyclohexane	0.590	0.617	-4.6	88	0.00
40 TM	Benzene	1.479	1.508	-2.0	89	0.00
41 T	Methacrylonitrile	0.124	0.109	12.1	71	0.00
42 TM	1,2-Dichloroethane	0.417	0.414	0.7	88	0.00
43 T	Isopropyl Acetate	0.442	0.423	4.3	83	0.00
44 TM	Trichloroethene	0.377	0.395	-4.8	94	0.00
45 C	1,2-Dichloropropane	0.350	0.358	-2.3#	90	0.00
46 T	Dibromomethane	0.202	0.206	-2.0	92	0.00
47 T	Bromodichloromethane	0.523	0.547	-4.6	92	0.00
48 T	Methyl methacrylate	0.203	0.207	-2.0	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041725\
 Data File : VY021900.D
 Acq On : 17 Apr 2025 09:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 18 03:52:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.234	1.231	0.2	84	0.00
51 T	4-Methyl-2-Pentanone	0.230	0.233	-1.3	86	0.00
52 CM	Toluene	0.923	0.982	-6.4#	92	0.00
53 T	t-1,3-Dichloropropene	0.465	0.478	-2.8	87	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.559	-2.0	88	0.00
55 T	1,1,2-Trichloroethane	0.260	0.273	-5.0	93	0.00
56 T	Ethyl methacrylate	0.332	0.360	-8.4	87	0.00
57 T	1,3-Dichloropropane	0.449	0.463	-3.1	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.162	0.185	-14.2	91	0.00
59 T	2-Hexanone	0.153	0.157	-2.6	85	0.00
60 T	Dibromochloromethane	0.355	0.377	-6.2	92	0.00
61 T	1,2-Dibromoethane	0.244	0.254	-4.1	89	0.00
62 S	4-Bromofluorobenzene	0.417	0.428	-2.6	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.455	0.508	-11.6	102	0.00
65 PM	Chlorobenzene	1.146	1.180	-3.0	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.428	-5.2	94	0.00
67 C	Ethyl Benzene	1.965	2.085	-6.1#	91	0.00
68 T	m/p-Xylenes	0.753	0.802	-6.5	92	0.00
69 T	o-Xylene	0.695	0.747	-7.5	91	0.00
70 T	Styrene	1.158	1.278	-10.4	93	0.00
71 P	Bromoform	0.232	0.243	-4.7	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.714	3.873	-4.3	93	0.00
74 T	N-amyl acetate	0.872	0.836	4.1	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.668	0.629	5.8	88	0.00
76 T	1,2,3-Trichloropropane	0.496	0.443	10.7	81	0.00
77 T	Bromobenzene	0.892	0.914	-2.5	94	0.00
78 T	n-propylbenzene	4.465	4.696	-5.2	93	0.00
79 T	2-Chlorotoluene	2.593	2.666	-2.8	92	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.211	-5.8	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.205	4.7	85	0.00
82 T	4-Chlorotoluene	2.704	2.764	-2.2	92	0.00
83 T	tert-Butylbenzene	2.703	2.907	-7.5	95	0.00
84 T	1,2,4-Trimethylbenzene	3.000	3.207	-6.9	94	0.00
85 T	sec-Butylbenzene	3.951	4.237	-7.2	95	0.00
86 T	p-Isopropyltoluene	3.265	3.584	-9.8	96	0.00
87 T	1,3-Dichlorobenzene	1.757	1.826	-3.9	96	0.00
88 T	1,4-Dichlorobenzene	1.736	1.797	-3.5	94	0.00
89 T	n-Butylbenzene	3.020	3.317	-9.8	96	0.00
90 T	Hexachloroethane	0.716	0.733	-2.4	95	0.00
91 T	1,2-Dichlorobenzene	1.529	1.600	-4.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.101	1.9	91	0.00
93 T	1,2,4-Trichlorobenzene	0.829	0.953	-15.0	99	0.00
94 T	Hexachlorobutadiene	0.518	0.591	-14.1	103	0.00
95 T	Naphthalene	1.375	1.597	-16.1	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041725\
Data File : VY021900.D
Acq On : 17 Apr 2025 09:00
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 18 03:52:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.707	0.829	-17.3	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6