

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041822\
 Data File : VY008373.D
 Acq On : 18 Apr 2022 15:09
 Operator : KP/MD
 Sample : N2430-01
 Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 E13-EB1-041322-0-2

Quant Time: Apr 19 06:29:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	97696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	178081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	170109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71590	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	56342	61.396	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.800%
35) Dibromofluoromethane	7.722	113	57693	56.091	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.180%
50) Toluene-d8	10.179	98	218519	62.087	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	124.180%
62) 4-Bromofluorobenzene	12.483	95	72199	54.540	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041822\
Data File : VY008373.D
Acq On : 18 Apr 2022 15:09
Operator : KP/MD
Sample : N2430-01
Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
E13-EB1-041322-0-2

Quant Time: Apr 19 06:29:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

