

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041924\
 Data File : VY017994.D
 Acq On : 19 Apr 2024 13:23
 Operator : SY/MD
 Sample : P2147-13
 Misc : 4.64g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-9

Quant Time: Apr 19 23:24:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	366943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	640595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	560820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	200443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	167490	52.746	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.500%	
35) Dibromofluoromethane	7.722	113	201622	51.017	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.040%	
50) Toluene-d8	10.185	98	730376	48.704	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.400%	
62) 4-Bromofluorobenzene	12.489	95	196500	42.309	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 84.620%	
Target Compounds						
20) Methylene Chloride	4.722	84	54248	12.762	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041924\
Data File : VY017994.D
Acq On : 19 Apr 2024 13:23
Operator : SY/MD
Sample : P2147-13
Misc : 4.64g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAB-9

Quant Time: Apr 19 23:24:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

