

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041924\
 Data File : VY018003.D
 Acq On : 19 Apr 2024 16:58
 Operator : SY/MD
 Sample : P2194-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-01-041824

Quant Time: Apr 19 23:27:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	318957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	544009	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	514639	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	221141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	147574	53.466	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 106.940%		
35) Dibromofluoromethane	7.716	113	174370	51.955	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.900%		
50) Toluene-d8	10.179	98	627723	49.291	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 98.580%		
62) 4-Bromofluorobenzene	12.483	95	192487	48.804	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 97.600%		
Target Compounds						
						Qvalue
16) Acetone	3.948	43	3933	7.212	ug/l #	79
20) Methylene Chloride	4.710	84	13356	3.615	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041924\
Data File : VY018003.D
Acq On : 19 Apr 2024 16:58
Operator : SY/MD
Sample : P2194-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-01-041824

Quant Time: Apr 19 23:27:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

