

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
 Data File : VY013364.D
 Acq On : 20 Apr 2023 13:30
 Operator : KP/MD
 Sample : VY0420SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 21 02:14:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	196545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	218711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	323121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	104792	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	103513	43.824	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.640%
35) Dibromofluoromethane	7.716	113	90750	64.093	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.180%
50) Toluene-d8	10.179	98	268040	50.859	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.720%
62) 4-Bromofluorobenzene	12.477	95	86874	50.733	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.460%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.900	85	33333	18.844	ug/l	94
3) Chloromethane	2.107	50	49438	22.308	ug/l	98
4) Vinyl Chloride	2.247	62	56974	27.116	ug/l	98
5) Bromomethane	2.650	94	38216	23.718	ug/l	99
6) Chloroethane	2.784	64	30637	21.621	ug/l	99
7) Trichlorofluoromethane	3.119	101	88405	27.033	ug/l	97
8) Diethyl Ether	3.521	74	12006	10.105	ug/l	66
9) 1,1,2-Trichlorotrifluo...	3.887	101	20428	10.630	ug/l #	78
10) Methyl Iodide	4.082	142	26739	11.343	ug/l #	65
11) Tert butyl alcohol	4.979	59	21172	31.762	ug/l	98
12) 1,1-Dichloroethene	3.869	96	16658	9.007	ug/l	94
13) Acrolein	3.729	56	6772	23.339	ug/l	92
14) Allyl chloride	4.473	41	40299	11.407	ug/l #	64
15) Acrylonitrile	5.155	53	35518	52.021	ug/l	96
16) Acetone	3.948	43	30879	38.022	ug/l #	80
17) Carbon Disulfide	4.192	76	64042	10.518	ug/l #	95
18) Methyl Acetate	4.473	43	27938	10.637	ug/l #	90
19) Methyl tert-butyl Ether	5.216	73	91338	16.486	ug/l #	88
20) Methylene Chloride	4.710	84	63296	27.593	ug/l	93
21) trans-1,2-Dichloroethene	5.210	96	31168	14.763	ug/l	91
22) Diisopropyl ether	6.112	45	169842	24.393	ug/l	90
23) Vinyl Acetate	6.058	43	482524	109.626	ug/l	99
24) 1,1-Dichloroethane	6.003	63	42186	10.847	ug/l #	91
25) 2-Butanone	6.984	43	77730	74.751	ug/l	95
26) 2,2-Dichloropropane	6.978	77	79530	22.792	ug/l	97
27) cis-1,2-Dichloroethene	6.978	96	50090	21.819	ug/l	95
28) Bromochloromethane	7.338	49	20542	13.252	ug/l #	56
29) Tetrahydrofuran	7.350	42	41053	65.003	ug/l #	77
30) Chloroform	7.502	83	80802	20.530	ug/l	99
31) Cyclohexane	7.777	56	52905	14.813	ug/l #	79
32) 1,1,1-Trichloroethane	7.698	97	80664	22.655	ug/l	99
36) 1,1-Dichloropropene	7.917	75	41652	19.418	ug/l	96
37) Ethyl Acetate	7.070	43	31353	18.133	ug/l	96
38) Carbon Tetrachloride	7.893	117	72362	30.509	ug/l	99
39) Methylcyclohexane	9.191	83	64016	25.672	ug/l	97
40) Benzene	8.161	78	133476	21.500	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
 Data File : VY013364.D
 Acq On : 20 Apr 2023 13:30
 Operator : KP/MD
 Sample : VY0420SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 21 02:14:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.301	41	12983	18.099	ug/l #	100
42) 1,2-Dichloroethane	8.234	62	53043	25.200	ug/l	91
43) Isopropyl Acetate	8.271	43	37799	13.161	ug/l #	82
44) Trichloroethene	8.935	130	69750	43.613	ug/l	94
45) 1,2-Dichloropropane	9.216	63	27903	17.316	ug/l	97
46) Dibromomethane	9.307	93	28261	29.141	ug/l	90
47) Bromodichloromethane	9.490	83	53313	24.047	ug/l	96
48) Methyl methacrylate	9.289	41	24201	18.345	ug/l #	82
49) 1,4-Dioxane	9.307	88	8645	657.935	ug/l #	78
51) 4-Methyl-2-Pentanone	10.075	43	143630	91.893	ug/l	97
52) Toluene	10.240	92	72832	19.428	ug/l	99
53) t-1,3-Dichloropropene	10.459	75	90228	38.862	ug/l	100
54) cis-1,3-Dichloropropene	9.929	75	52446	20.601	ug/l #	84
55) 1,1,2-Trichloroethane	10.642	97	27912	21.828	ug/l	95
56) Ethyl methacrylate	10.508	69	55640	31.779	ug/l	91
57) 1,3-Dichloropropane	10.788	76	39215	18.031	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.783	63	55773	76.291	ug/l #	79
59) 2-Hexanone	10.831	43	75819	66.714	ug/l	99
60) Dibromochloromethane	10.990	129	48954	30.532	ug/l	98
61) 1,2-Dibromoethane	11.087	107	50362	40.446	ug/l	100
64) Tetrachloroethene	10.715	164	39772	17.447	ug/l	97
65) Chlorobenzene	11.514	112	150288	22.234	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.587	131	58603	23.103	ug/l	96
67) Ethyl Benzene	11.593	91	251926	21.399	ug/l	97
68) m/p-Xylenes	11.703	106	190276	42.243	ug/l	100
69) o-Xylene	12.026	106	54755	13.125	ug/l	97
70) Styrene	12.044	104	89081	12.643	ug/l	95
71) Bromoform	12.209	173	26167	14.890	ug/l #	96
73) Isopropylbenzene	12.325	105	151568	21.086	ug/l	100
74) N-amyl acetate	12.142	43	26794	10.448	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.581	83	21202	11.940	ug/l #	94
76) 1,2,3-Trichloropropane	12.630	75	38974	29.847	ug/l #	100
77) Bromobenzene	12.605	156	41434	23.051	ug/l	76
78) n-propylbenzene	12.672	91	177941	20.044	ug/l	94
79) 2-Chlorotoluene	12.758	91	97678	19.547	ug/l	92
80) 1,3,5-Trimethylbenzene	12.813	105	136970	22.749	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.374	75	8775	13.947	ug/l #	65
82) 4-Chlorotoluene	12.849	91	113365	21.519	ug/l	96
83) tert-Butylbenzene	13.075	119	122288	23.670	ug/l	98
84) 1,2,4-Trimethylbenzene	13.117	105	135739	23.165	ug/l	98
85) sec-Butylbenzene	13.251	105	182687	23.646	ug/l	93
86) p-Isopropyltoluene	13.367	119	148644	23.447	ug/l	94
87) 1,3-Dichlorobenzene	13.367	146	78731	22.722	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	79592	22.731	ug/l	100
89) n-Butylbenzene	13.696	91	159428	26.293	ug/l	95
90) Hexachloroethane	13.959	117	28714	21.722	ug/l	73
91) 1,2-Dichlorobenzene	13.739	146	92783	29.427	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	7706	23.897	ug/l	94
93) 1,2,4-Trichlorobenzene	15.007	180	58245	31.081	ug/l	99
94) Hexachlorobutadiene	15.111	225	35489	31.982	ug/l	98
95) Naphthalene	15.233	128	153688	34.507	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	47612	27.888	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
Data File : VY013364.D
Acq On : 20 Apr 2023 13:30
Operator : KP/MD
Sample : VY0420SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :


Quant Time: Apr 21 02:14:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 21 02:14:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 196545
Ion Ratio Lower Upper
168 100
99 54.9 41.7 62.5



#2
Dichlorodifluoromethane
Concen: 18.844 ug/l
RT: 1.900 min Scan# 13
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 85 Resp: 33333
Ion Ratio Lower Upper
85 100
87 36.6 16.7 50.0









#3
Chloromethane
Concen: 22.308 ug/l
RT: 2.107 min Scan# 41
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 50 Resp: 49438
Ion Ratio Lower Upper
50 100
52 33.8 26.0 39.0



#4
Vinyl Chloride
Concen: 27.116 ug/l
RT: 2.247 min Scan# 70
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 62 Resp: 56974
Ion Ratio Lower Upper
62 100
64 33.4 25.6 38.4









#5
Bromomethane
Concen: 23.718 ug/l
RT: 2.650 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 94 Resp: 38216
Ion Ratio Lower Upper
94 100
96 99.4 80.3 120.5



#6
Chloroethane
Concen: 21.621 ug/l
RT: 2.784 min Scan# 158
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 64 Resp: 30637
Ion Ratio Lower Upper
64 100
66 32.6 25.7 38.5









#7
Trichlorofluoromethane
Concen: 27.033 ug/l
RT: 3.119 min Scan# 214
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 88405
Ion Ratio Lower Upper
101 100
103 67.1 51.5 77.3



#8
Diethyl Ether
Concen: 10.105 ug/l
RT: 3.521 min Scan# 279
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 74 Resp: 12006
Ion Ratio Lower Upper
74 100
45 69.5 52.1 156.4








#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.630 ug/l
RT: 3.887 min Scan# 311
Delta R.T. -0.012 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:101 Resp: 20428
Ion Ratio Lower Upper
101 100
85 54.4 35.3 52.9#
151 101.8 63.1 94.7#



#10
Methyl Iodide
Concen: 11.343 ug/l
RT: 4.082 min Scan# 371
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:142 Resp: 26739
Ion Ratio Lower Upper
142 100
127 73.7 35.4 53.2#
141 15.4 11.3 16.9





#11

Tert butyl alcohol

Concen: 31.762 ug/l

RT: 4.979 min Scan# 517

Delta R.T. 0.007 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 59 Resp: 21172

Ion Ratio Lower Upper

59 100

57 10.4 7.7 11.5



#12

1,1-Dichloroethene

Concen: 9.007 ug/l

RT: 3.869 min Scan# 336

Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30




Tgt Ion: 96 Resp: 16658


Ion Ratio Lower Upper

96 100

61 173.7 131.7 197.5


98 68.1 52.0 78.0








#13
Acrolein
Concen: 23.339 ug/l
RT: 3.729 min Scan# 313
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 56 Resp: 6772
Ion Ratio Lower Upper
56 100
55 77.9 56.8 85.2



#14
Allyl chloride
Concen: 11.407 ug/l
RT: 4.473 min Scan# 435
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 41 Resp: 40299
Ion Ratio Lower Upper
41 100
39 103.7 53.6 80.4#
76 44.5 27.2 40.8#








#15
Acrylonitrile
Concen: 52.021 ug/l
RT: 5.155 min Scan# 54
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 53 Resp: 35518
Ion Ratio Lower Upper
53 100
52 82.2 66.3 99.5
51 43.0 28.8 43.2



#16
Acetone
Concen: 38.022 ug/l
RT: 3.948 min Scan# 349
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 43 Resp: 30879
Ion Ratio Lower Upper
43 100
58 19.5 24.2 36.2#









#17
Carbon Disulfide
Concen: 10.518 ug/l
RT: 4.192 min Scan# 389
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 76 Resp: 64042
Ion Ratio Lower Upper
76 100
78 10.8 7.1 10.7#



#18
Methyl Acetate
Concen: 10.637 ug/l
RT: 4.473 min Scan# 435
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 43 Resp: 27938
Ion Ratio Lower Upper
43 100
74 26.5 17.4 26.0#









#19
Methyl tert-butyl Ether
Concen: 16.486 ug/l
RT: 5.216 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 73 Resp: 91338
Ion Ratio Lower Upper
73 100
57 16.9 18.3 27.5#



#20
Methylene Chloride
Concen: 27.593 ug/l
RT: 4.710 min Scan# 474
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 84 Resp: 63296
Ion Ratio Lower Upper
84 100
49 118.2 102.6 154.0
51 34.2 31.9 47.9
86 63.2 51.0 76.6









#21
trans-1,2-Dichloroethene
Concen: 14.763 ug/l
RT: 5.210 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 96 Resp: 31168
Ion Ratio Lower Upper
96 100
61 125.1 111.4 167.0
98 61.6 50.8 76.2



#22
Diisopropyl ether
Concen: 24.393 ug/l
RT: 6.112 min Scan# 704
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 45 Resp: 169842
Ion Ratio Lower Upper
45 100
43 65.6 45.0 67.6
87 29.6 20.3 30.5
59 11.0 8.8 13.2









#23
 Vinyl Acetate
 Concen: 109.626 ug/l
 RT: 6.058 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 43 Resp: 482524
 Ion Ratio Lower Upper
 43 100
 86 10.3 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 10.847 ug/l
 RT: 6.003 min Scan# 686
 Delta R.T. -0.012 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30

Tgt Ion: 63 Resp: 42186
 Ion Ratio Lower Upper
 63 100
 98 10.5 3.7 11.1
 100 7.9 2.3 6.8#





#25

2-Butanone

Concen: 74.751 ug/l

RT: 6.984 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 43 Resp: 77730

Ion Ratio Lower Upper

43 100

72 28.9 21.0 31.4



#26

2,2-Dichloropropane

Concen: 22.792 ug/l

RT: 6.978 min Scan# 846

Delta R.T. -0.000 min

Lab File: VY013364.D




Acq: 20 Apr 2023 13:30


Tgt Ion: 77 Resp: 79530

Ion Ratio Lower Upper

77 100


97 21.8 11.7 35.0








#27
 cis-1,2-Dichloroethene
 Concen: 21.819 ug/l
 RT: 6.978 min Scan# 846
 Delta R.T. -0.000 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion: 96 Resp: 50090




Ion	Ratio	Lower	Upper
96	100		
61	138.4	0.0	295.8
98	64.6	0.0	130.2




#28
 Bromochloromethane
 Concen: 13.252 ug/l
 RT: 7.338 min Scan# 905
 Delta R.T. 0.012 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30

Tgt Ion: 49 Resp: 20542

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.2
130	114.4	60.5	90.7#









#29
Tetrahydrofuran
Concen: 65.003 ug/l
RT: 7.350 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 42 Resp: 41053
Ion Ratio Lower Upper
42 100
72 53.7 32.3 48.5#
71 52.3 30.3 45.5#



#30
Chloroform
Concen: 20.530 ug/l
RT: 7.502 min Scan# 932
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 83 Resp: 80802
Ion Ratio Lower Upper
83 100
85 65.2 51.4 77.0








#31
Cyclohexane
Concen: 14.813 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 56 Resp: 52905
Ion Ratio Lower Upper
56 100
69 45.1 25.7 38.5#
84 98.8 64.3 96.5#




#32
1,1,1-Trichloroethane
Concen: 22.655 ug/l
RT: 7.698 min Scan# 964
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30



Tgt Ion: 97 Resp: 80664
Ion Ratio Lower Upper
97 100
99 64.3 51.0 76.6
61 41.6 34.5 51.7









#33
1,2-Dichloroethane-d4
Concen: 43.824 ug/l
RT: 8.143 min Scan# 1036
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 65 Resp: 103513
Ion Ratio Lower Upper
65 100
67 46.4 0.0 104.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion:114 Resp: 218711
Ion Ratio Lower Upper
114 100
63 17.1 0.0 37.4
88 13.9 0.0 28.6








#35
Dibromofluoromethane
Concen: 64.093 ug/l
RT: 7.716 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:113 Resp: 90750
Ion Ratio Lower Upper
113 100
111 103.4 82.2 123.2
192 24.2 16.8 25.2



#36
1,1-Dichloropropene
Concen: 19.418 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 75 Resp: 41652
Ion Ratio Lower Upper
75 100
110 40.1 17.8 53.3
77 31.2 25.0 37.4





#37

Ethyl Acetate

Concen: 18.133 ug/l

RT: 7.070 min Scan# 861

Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 43 Resp: 31353

Ion Ratio Lower Upper

43 100

61 14.7 10.7 16.1

70 11.6 7.8 11.8



#38

Carbon Tetrachloride

Concen: 30.509 ug/l

RT: 7.893 min Scan# 996

Delta R.T. -0.006 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30




Tgt Ion: 117 Resp: 72362


Ion Ratio Lower Upper

117 100

119 95.0 77.2 115.8

121 30.4 24.7 37.1





#39

Methylcyclohexane

Concen: 25.672 ug/l

RT: 9.191 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 64016

Ion Ratio Lower Upper

83 100

55 71.4 60.0 90.0

98 47.2 38.3 57.5



#40

Benzene

Concen: 21.500 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY013364.D




Acq: 20 Apr 2023 13:30


Tgt Ion: 78 Resp: 133476

Ion Ratio Lower Upper

78 100


77 23.7 19.4 29.0








#41
Methacrylonitrile
Concen: 18.099 ug/l
RT: 7.301 min Scan# 899
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:	41	Resp:	12983
Ion	Ratio	Lower	Upper
41	100		
39	73.6	0.0	0.0#
67	104.1	0.0	0.0#
52	46.5	0.0	0.0#



#42
1,2-Dichloroethane
Concen: 25.200 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion:	62	Resp:	53043
Ion	Ratio	Lower	Upper
62	100		
98	6.2	0.0	18.8








#43
Isopropyl Acetate
Concen: 13.161 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 37799
Ion Ratio Lower Upper
43 100
61 15.6 21.6 32.4#
87 15.9 9.9 14.9#



#44
Trichloroethene
Concen: 43.613 ug/l
RT: 8.935 min Scan# 1167
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:130 Resp: 69750
Ion Ratio Lower Upper
130 100
95 97.8 0.0 185.0





#45

1,2-Dichloropropane

Concen: 17.316 ug/l

RT: 9.216 min Scan# 1212

Delta R.T. 0.006 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 63 Resp: 27903

Ion Ratio Lower Upper

63 100

65 30.6 25.7 38.5



#46

Dibromomethane

Concen: 29.141 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.006 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30




Tgt Ion: 93 Resp: 28261


Ion Ratio Lower Upper

93 100

95 83.4 65.6 98.4

174 119.1 81.3 121.9





#47

Bromodichloromethane

Concen: 24.047 ug/l

RT: 9.490 min Scan# 1259

Delta R.T. -0.006 min

Lab File: VY013364.D





Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion: 83	Resp: 53313
Ion Ratio	Lower Upper
83 100	
85 65.8	50.2 75.4
127 7.5	6.3 9.5



#48

Methyl methacrylate

Concen: 18.345 ug/l




RT: 9.289 min Scan# 1225


Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Tgt Ion: 41	Resp: 24201
Ion Ratio	Lower Upper
41 100	
69 89.2	64.1 96.1
39 76.5	44.0 66.0





#49

1,4-Dioxane

Concen: 657.935 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 88 Resp: 8645

Ion Ratio Lower Upper

88 100

43 0.0 28.6 42.8#

58 70.7 55.2 82.8



#50

Toluene-d8

Concen: 50.859 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. -0.000 min

Lab File: VY013364.D




Acq: 20 Apr 2023 13:30


Tgt Ion: 98 Resp: 268040

Ion Ratio Lower Upper

98 100


100 63.8 50.9 76.3








#51
4-Methyl-2-Pentanone
Concen: 91.893 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 143630
Ion Ratio Lower Upper
43 100
58 38.8 29.8 44.6



#52
Toluene
Concen: 19.428 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 92 Resp: 72832
Ion Ratio Lower Upper
92 100
91 178.1 141.3 211.9









#53
t-1,3-Dichloropropene
Concen: 38.862 ug/l
RT: 10.459 min Scan# 1417
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 90228
Ion Ratio Lower Upper
75 100
77 31.6 25.4 38.2



#54
cis-1,3-Dichloropropene
Concen: 20.601 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 75 Resp: 52446
Ion Ratio Lower Upper
75 100
77 29.9 25.4 38.0
39 65.7 39.0 58.4#








#55
1,1,2-Trichloroethane
Concen: 21.828 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion: 97 Resp: 27912




Ion	Ratio	Lower	Upper
97	100		
83	80.7	69.0	103.6
85	50.4	44.9	67.3
99	62.6	50.8	76.2




#56
Ethyl methacrylate
Concen: 31.779 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 69 Resp: 55640

Ion	Ratio	Lower	Upper
69	100		
41	62.8	56.4	84.6
39	42.8	30.8	46.2









#57
1,3-Dichloropropane
Concen: 18.031 ug/l
RT: 10.788 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 76 Resp: 39215
Ion Ratio Lower Upper
76 100
78 33.6 26.2 39.4



#58
2-Chloroethyl Vinyl ether
Concen: 76.291 ug/l
RT: 9.783 min Scan# 1306
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 63 Resp: 55773
Ion Ratio Lower Upper
63 100
106 38.7 22.1 33.1#









#59
2-Hexanone
Concen: 66.714 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 75819
Ion Ratio Lower Upper
43 100
58 51.5 25.9 77.8



#60
Dibromochloromethane
Concen: 30.532 ug/l
RT: 10.990 min Scan# 1504
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:129 Resp: 48954
Ion Ratio Lower Upper
129 100
127 75.6 38.7 116.1









#61
1,2-Dibromoethane
Concen: 40.446 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:107 Resp: 50362
Ion Ratio Lower Upper
107 100
109 93.7 75.0 112.6



#62
4-Bromofluorobenzene
Concen: 50.733 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 95 Resp: 86874
Ion Ratio Lower Upper
95 100
174 106.9 0.0 176.6
176 104.0 0.0 170.2








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:117 Resp: 323121
Ion Ratio Lower Upper
117 100
82 53.2 45.4 68.0
119 32.3 25.6 38.4



#64
Tetrachloroethene
Concen: 17.447 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion:164 Resp: 39772
Ion Ratio Lower Upper
164 100
166 125.0 103.4 155.2
129 86.8 71.4 107.2
131 84.8 69.9 104.9








#65
Chlorobenzene
Concen: 22.234 ug/l
RT: 11.514 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:112 Resp: 150288
Ion Ratio Lower Upper
112 100
114 34.1 24.5 36.7



#66
1,1,1,2-Tetrachloroethane
Concen: 23.103 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:131 Resp: 58603
Ion Ratio Lower Upper
131 100
133 93.2 48.0 144.0
119 61.3 32.6 98.0





#67

Ethyl Benzene

Concen: 21.399 ug/l

RT: 11.593 min Scan# 1603

Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 251926

Ion Ratio Lower Upper

91 100

106 33.1 25.1 37.7



#68

m/p-Xylenes

Concen: 42.243 ug/l

RT: 11.703 min Scan# 1621

Delta R.T. -0.000 min

Lab File: VY013364.D




Acq: 20 Apr 2023 13:30


Tgt Ion: 106 Resp: 190276

Ion Ratio Lower Upper

106 100

91 199.2 159.8 239.8









#69
o-Xylene
Concen: 13.125 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. 0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:106 Resp: 54755
Ion Ratio Lower Upper
106 100
91 210.4 107.2 321.6



#70
Styrene
Concen: 12.643 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:104 Resp: 89081
Ion Ratio Lower Upper
104 100
78 55.4 39.2 58.8
103 55.4 43.7 65.5









#71
Bromoform
Concen: 14.890 ug/l
RT: 12.209 min Scan# 1703
Delta R.T. 0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:173 Resp: 26167
Ion Ratio Lower Upper
173 100
175 47.4 25.0 75.0
254 0.1 0.2 0.2#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:152 Resp: 104792
Ion Ratio Lower Upper
152 100
115 54.5 27.1 81.3
150 164.8 0.0 347.8









#73
Isopropylbenzene
Concen: 21.086 ug/l
RT: 12.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 151568
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.2



#74
N-ethyl acetate
Concen: 10.448 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 43 Resp: 26794
Ion Ratio Lower Upper
43 100
70 41.5 31.6 47.4
55 25.7 18.8 28.2
61 21.9 17.8 26.8









#75
1,1,2,2-Tetrachloroethane
Concen: 11.940 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 21202
Ion Ratio Lower Upper
83 100
131 18.6 5.1 15.2#
85 65.9 31.8 95.3



#76
1,2,3-Trichloropropane
Concen: 29.847 ug/l
RT: 12.630 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 75 Resp: 38974
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0








#77
Bromobenzene
Concen: 23.051 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:156 Resp: 41434
Ion Ratio Lower Upper
156 100
77 135.3 94.0 282.1
158 96.8 48.1 144.4



#78
n-propylbenzene
Concen: 20.044 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 91 Resp: 177941
Ion Ratio Lower Upper
91 100
120 26.1 11.6 34.8









#79
2-Chlorotoluene
Concen: 19.547 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 97678
Ion Ratio Lower Upper
91 100
126 39.5 17.4 52.2



#80
1,3,5-Trimethylbenzene
Concen: 22.749 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30


Tgt Ion: 105 Resp: 136970
Ion Ratio Lower Upper
105 100
120 48.0 24.8 74.4








#81
trans-1,4-Dichloro-2-butene
Concen: 13.947 ug/l
RT: 12.374 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 75 Resp: 8775
Ion Ratio Lower Upper
75 100
53 130.6 73.4 110.0#
89 62.0 36.6 55.0#



#82
4-Chlorotoluene
Concen: 21.519 ug/l
RT: 12.849 min Scan# 1809
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 91 Resp: 113365
Ion Ratio Lower Upper
91 100
126 37.5 17.7 53.1









#83
tert-Butylbenzene
Concen: 23.670 ug/l
RT: 13.075 min Scan# 119
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:119 Resp: 122288
Ion Ratio Lower Upper
119 100
91 67.2 32.8 98.3
134 26.2 13.6 40.8



#84
1,2,4-Trimethylbenzene
Concen: 23.165 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:105 Resp: 135739
Ion Ratio Lower Upper
105 100
120 44.7 23.0 69.0









#85
 sec-Butylbenzene
 Concen: 23.646 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. -0.000 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:105 Resp: 182687
 Ion Ratio Lower Upper
 105 100
 134 24.1 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 23.447 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. -0.000 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30

Tgt Ion:119 Resp: 148644
 Ion Ratio Lower Upper
 119 100
 134 28.5 13.2 39.6
 91 27.5 11.8 35.3









#87
1,3-Dichlorobenzene
Concen: 22.722 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:146 Resp: 78731
Ion Ratio Lower Upper
146 100
111 37.0 19.1 57.5
148 63.1 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 22.731 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:146 Resp: 79592
Ion Ratio Lower Upper
146 100
111 37.3 18.6 55.8
148 63.4 31.5 94.5









#89
n-Butylbenzene
Concen: 26.293 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 159428
Ion Ratio Lower Upper
91 100
92 50.4 26.4 79.0
134 30.6 13.0 38.9



#90
Hexachloroethane
Concen: 21.722 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion: 117 Resp: 28714
Ion Ratio Lower Upper
117 100
201 102.2 39.4 118.1





#91

1,2-Dichlorobenzene

Concen: 29.427 ug/l

RT: 13.739 min Scan# 1955

Delta R.T. -0.000 min

Lab File: VY013364.D





Acq: 20 Apr 2023 13:30

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion:146	Resp:	92783
Ion	Ratio	Lower Upper
146	100	
111	37.8	19.7 59.1
148	63.7	31.9 95.5



#92

1,2-Dibromo-3-Chloropropane

Concen: 23.897 ug/l




RT: 14.355 min Scan# 2056


Delta R.T. -0.000 min

Lab File: VY013364.D

Acq: 20 Apr 2023 13:30

Tgt Ion: 75	Resp:	7706
Ion	Ratio	Lower Upper
75	100	
155	79.2	43.8 131.4
157	104.0	54.1 162.3









#93
 1,2,4-Trichlorobenzene
 Concen: 31.081 ug/l
 RT: 15.007 min Scan# 2163
 Delta R.T. 0.000 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:180 Resp: 58245
 Ion Ratio Lower Upper
 180 100
 182 94.6 46.7 140.1
 145 29.7 14.6 43.8



#94
 Hexachlorobutadiene
 Concen: 31.982 ug/l
 RT: 15.111 min Scan# 2180
 Delta R.T. -0.000 min
 Lab File: VY013364.D
 Acq: 20 Apr 2023 13:30


Tgt Ion:225 Resp: 35489
 Ion Ratio Lower Upper
 225 100
 223 62.6 31.1 93.3
 227 66.8 32.4 97.0








#95
Naphthalene
Concen: 34.507 ug/l
RT: 15.233 min Scan# 21
Delta R.T. -0.006 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion:128 Resp: 153688
Ion Ratio Lower Upper
128 100
127 13.2 10.3 15.5
129 10.6 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 27.888 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. -0.000 min
Lab File: VY013364.D
Acq: 20 Apr 2023 13:30

Tgt Ion:180 Resp: 47612
Ion Ratio Lower Upper
180 100
182 92.0 47.4 142.3
145 29.9 15.3 45.8

