

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
 Data File : VY013365.D
 Acq On : 20 Apr 2023 14:03
 Operator : KP/MD
 Sample : 02371-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-26-9.0-9.5

Quant Time: Apr 21 02:14:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	81341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	96367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	209420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	59822	61.197	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.400%	
35) Dibromofluoromethane	7.710	113	35663	57.164	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	114.320%	
50) Toluene-d8	10.167	98	137794	59.339	ug/l	-0.01
Spiked Amount	50.000	Range 49 - 140	Recovery	=	118.680%	
62) 4-Bromofluorobenzene	12.483	95	71892	95.284	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	190.560%#	
Target Compounds						
16) Acetone	3.973	43	3108	9.247	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
Data File : VY013365.D
Acq On : 20 Apr 2023 14:03
Operator : KP/MD
Sample : 02371-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-26-9.0-9.5

Quant Time: Apr 21 02:14:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

