

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
 Data File : VY013375.D
 Acq On : 20 Apr 2023 17:57
 Operator : KP/MD
 Sample : 02406-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-12-6.0-6.5

Quant Time: Apr 21 02:16:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	150006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	286302	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	296078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	74729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	85348	47.344	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.680%
35) Dibromofluoromethane	7.722	113	56776	30.632	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.260%
50) Toluene-d8	10.185	98	276906	40.137	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	80.280%
62) 4-Bromofluorobenzene	12.483	95	113472	50.621	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.240%
Target Compounds						
16) Acetone	3.960	43	5641	9.101	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
Data File : VY013375.D
Acq On : 20 Apr 2023 17:57
Operator : KP/MD
Sample : 02406-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-12-6.0-6.5

Quant Time: Apr 21 02:16:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

