



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042023\
Data File : VY013360.D
Acq On : 20 Apr 2023 10:15
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Title : SW846 8260
Last Update : Thu Apr 13 05:55:16 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	22608	PASS
75	95	30	60	51.6	59603	PASS
95	95	100	100	100.0	115437	PASS
96	95	5	9	6.5	7517	PASS
173	174	0.00	2	1.4	1489	PASS
174	95	50	100	92.3	106603	PASS
175	174	5	9	7.9	8474	PASS
176	174	95	101	96.8	103237	PASS
177	176	5	9	7.0	7224	PASS