

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008450.D
 Acq On : 21 Apr 2022 18:59
 Operator : KP/MD
 Sample : N2511-03
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1528

Quant Time: Apr 22 05:27:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	1321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	147	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	710	89.528	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	179.060%#
35) Dibromofluoromethane	7.716	113	406	47.622	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.240%
50) Toluene-d8	10.179	98	1599	47.191	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.483	95	516	44.948	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
Data File : VY008450.D
Acq On : 21 Apr 2022 18:59
Operator : KP/MD
Sample : N2511-03
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1528

Quant Time: Apr 22 05:27:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

