

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008434.D
 Acq On : 21 Apr 2022 10:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Apr 22 05:32:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	134294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	226529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	208088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	101505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	57817	36.590	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	73.180%
35) Dibromofluoromethane	7.716	113	61404	42.001	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.000%
50) Toluene-d8	10.179	98	209381	36.547	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.100%
62) 4-Bromofluorobenzene	12.483	95	78579	39.916	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.840%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.906	85	36415	29.860	ug/l	98
3) Chloromethane	2.113	50	54722	28.516	ug/l	92
4) Vinyl Chloride	2.253	62	77042	30.897	ug/l	99
5) Bromomethane	2.650	94	68978	29.751	ug/l	96
6) Chloroethane	2.796	64	56329	33.956	ug/l	98
7) Trichlorofluoromethane	3.125	101	94426	36.555	ug/l	97
8) Diethyl Ether	3.527	74	34389	38.681	ug/l	70
9) 1,1,2-Trichlorotrifluo...	3.905	101	62005	40.324	ug/l	93
10) Methyl Iodide	4.094	142	63211	31.837	ug/l	92
11) Tert butyl alcohol	4.954	59	36448	225.352	ug/l #	95
12) 1,1-Dichloroethene	3.875	96	53089	36.570	ug/l #	80
13) Acrolein	3.729	56	20043	119.139	ug/l	96
14) Allyl chloride	4.479	41	74638	39.390	ug/l #	91
15) Acrylonitrile	5.161	53	96051	211.303	ug/l	97
16) Acetone	3.948	43	80418	218.501	ug/l	91
17) Carbon Disulfide	4.198	76	111723	25.456	ug/l	99
18) Methyl Acetate	4.472	43	38611	37.408	ug/l #	82
19) Methyl tert-butyl Ether	5.222	73	194420	42.326	ug/l	93
20) Methylene Chloride	4.716	84	72096	35.929	ug/l #	82
21) trans-1,2-Dichloroethene	5.222	96	63290	35.740	ug/l	83
22) Diisopropyl ether	6.118	45	194145	42.428	ug/l #	93
23) Vinyl Acetate	6.058	43	657372	209.543	ug/l #	89
24) 1,1-Dichloroethane	6.021	63	118640	41.653	ug/l	98
25) 2-Butanone	6.984	43	120860	212.262	ug/l #	79
26) 2,2-Dichloropropane	6.984	77	115894	43.931	ug/l	94
27) cis-1,2-Dichloroethene	6.990	96	83750	41.346	ug/l	82
28) Bromochloromethane	7.332	49	51810	45.094	ug/l #	74
29) Tetrahydrofuran	7.344	42	76162	207.677	ug/l #	75
30) Chloroform	7.508	83	135079	43.167	ug/l	98
31) Cyclohexane	7.783	56	86337	33.236	ug/l	86
32) 1,1,1-Trichloroethane	7.704	97	117100	42.737	ug/l	95
36) 1,1-Dichloropropene	7.917	75	89482	38.478	ug/l	96
37) Ethyl Acetate	7.076	43	53283	41.429	ug/l #	92
38) Carbon Tetrachloride	7.899	117	101876	42.164	ug/l	99
39) Methylcyclohexane	9.185	83	109701	36.842	ug/l #	84
40) Benzene	8.161	78	273339	40.288	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008434.D
 Acq On : 21 Apr 2022 10:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022

Quant Time: Apr 22 05:32:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	26386m	42.306	ug/l	04/22/2022
42) 1,2-Dichloroethane	8.240	62	85344	41.781	ug/l	Supervised By :Semsettin Yesilyurt
43) Isopropyl Acetate	8.271	43	103658	43.824	ug/l #	87
44) Trichloroethene	8.941	130	78256	41.867	ug/l	97
45) 1,2-Dichloropropane	9.215	63	72322	44.390	ug/l	96
46) Dibromomethane	9.307	93	47912	41.975	ug/l	94
47) Bromodichloromethane	9.496	83	109053	45.006	ug/l	100
48) Methyl methacrylate	9.295	41	43501	43.253	ug/l #	04/22/2022
49) 1,4-Dioxane	9.301	88	12446	864.691	ug/l #	80
51) 4-Methyl-2-Pentanone	10.069	43	280569	225.483	ug/l #	85
52) Toluene	10.246	92	187719	41.838	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	114669	44.129	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	123122	42.758	ug/l #	82
55) 1,1,2-Trichloroethane	10.648	97	69308	44.488	ug/l	97
56) Ethyl methacrylate	10.508	69	90390	44.935	ug/l #	75
57) 1,3-Dichloropropane	10.794	76	111169	43.713	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	152829	138.628	ug/l	90
59) 2-Hexanone	10.831	43	196680	229.694	ug/l	80
60) Dibromochloromethane	10.983	129	83237	46.504	ug/l	100
61) 1,2-Dibromoethane	11.087	107	65707	42.643	ug/l	98
64) Tetrachloroethene	10.721	164	63993	41.469	ug/l	96
65) Chlorobenzene	11.514	112	212514	45.167	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.593	131	83561	48.103	ug/l	96
67) Ethyl Benzene	11.593	91	372301	45.593	ug/l	98
68) m/p-Xylenes	11.703	106	289679	88.564	ug/l	93
69) o-Xylene	12.032	106	140842	44.726	ug/l	91
70) Styrene	12.044	104	242205	45.642	ug/l	94
71) Bromoform	12.209	173	52053	47.815	ug/l #	98
73) Isopropylbenzene	12.331	105	375602	48.805	ug/l	100
74) N-amyl acetate	12.142	43	93807	49.331	ug/l #	83
75) 1,1,2,2-Tetrachloroethane	12.587	83	86899	50.536	ug/l	99
76) 1,2,3-Trichloropropane	12.636	75	53294m	40.570	ug/l	
77) Bromobenzene	12.611	156	88033	46.903	ug/l	95
78) n-propylbenzene	12.672	91	443034	47.993	ug/l	98
79) 2-Chlorotoluene	12.758	91	247878	47.710	ug/l	97
80) 1,3,5-Trimethylbenzene	12.812	105	307531	48.141	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.380	75	28713	50.707	ug/l #	82
82) 4-Chlorotoluene	12.855	91	256263	47.598	ug/l	97
83) tert-Butylbenzene	13.075	119	277421	50.027	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	309848	48.885	ug/l	98
85) sec-Butylbenzene	13.257	105	417922	49.937	ug/l	99
86) p-Isopropyltoluene	13.373	119	351739	50.185	ug/l	97
87) 1,3-Dichlorobenzene	13.367	146	178613	47.532	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	176731	47.478	ug/l	98
89) n-Butylbenzene	13.696	91	317288	49.877	ug/l	99
90) Hexachloroethane	13.965	117	67117	53.264	ug/l	81
91) 1,2-Dichlorobenzene	13.739	146	159779	47.820	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.361	75	12300	47.597	ug/l	74
93) 1,2,4-Trichlorobenzene	15.007	180	91522	49.209	ug/l	99
94) Hexachlorobutadiene	15.111	225	46142	50.847	ug/l	99
95) Naphthalene	15.239	128	204605	48.527	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	79646	48.926	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008434.D
 Acq On : 21 Apr 2022 10:40
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022

Quant Time: Apr 22 05:32:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

04/22/2022
 Supervised By :Semsettin
 Yesilyurt

04/22/2022


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
Data File : VY008434.D
Acq On : 21 Apr 2022 10:40
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/soil
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_Y
Client Sample Id :
VSTDCCC050

Quant Time: Apr 22 05:32:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 134294
Ion Ratio Lower Upper
168 100
99 54.6 46.9 70.3


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#2
Dichlorodifluoromethane
Concen: 29.860 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 85 Resp: 36415
Ion Ratio Lower Upper
85 100
87 33.2 16.1 48.2





#3
Chloromethane
Concen: 28.516 ug/l
RT: 2.113 min Scan# 48
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
Client SampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 5472
Ion Ratio Lower Upper
50 100
52 34.9 24.6 37.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#4
Vinyl Chloride
Concen: 30.897 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 62 Resp: 77042
Ion Ratio Lower Upper
62 100
64 31.8 25.0 37.6





#5
Bromomethane
Concen: 29.751 ug/l
RT: 2.650 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 68978
Ion Ratio Lower Upper
94 100
96 95.6 73.5 110.3


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#6
Chloroethane
Concen: 33.956 ug/l
RT: 2.796 min Scan# 160
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 64 Resp: 56329
Ion Ratio Lower Upper
64 100
66 32.6 25.4 38.0





#7
Trichlorofluoromethane
Concen: 36.555 ug/l
RT: 3.125 min Scan# 214
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 101 Resp: 94420
Ion Ratio Lower Upper
101 100
103 62.1 48.0 72.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#8
Diethyl Ether
Concen: 38.681 ug/l
RT: 3.527 min Scan# 280
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 74 Resp: 34389
Ion Ratio Lower Upper
74 100
45 79.0 55.1 165.5







#9
1,1,2-Trichlorotrifluoroethane
Concen: 40.324 ug/l
RT: 3.905 min Scan# 341
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 6200
Ion Ratio Lower Upper
101 100
85 45.3 40.6 60.8
151 75.5 64.9 97.3



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#10
Methyl Iodide
Concen: 31.837 ug/l
RT: 4.094 min Scan# 373
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:142 Resp: 63211
Ion Ratio Lower Upper
142 100
127 45.1 41.5 62.3
141 14.2 11.9 17.9







#11
Tert butyl alcohol
Concen: 225.352 ug/l
RT: 4.954 min Scan# 513
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 59 Resp: 36448
Ion Ratio Lower Upper
59 100
57 9.7 6.4 9.6



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#12
1,1-Dichloroethene
Concen: 36.570 ug/l
RT: 3.875 min Scan# 337
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 96 Resp: 53089
Ion Ratio Lower Upper
96 100
61 145.3 145.8 218.6#
98 62.5 52.5 78.7





#13
Acrolein
Concen: 119.139 ug/l
RT: 3.729 min Scan# 313
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 2004
Ion Ratio Lower Upper
56 100
55 69.1 57.6 86.4


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#14
Allyl chloride
Concen: 39.390 ug/l
RT: 4.479 min Scan# 436
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 41 Resp: 74638
Ion Ratio Lower Upper
41 100
39 74.4 58.8 88.2
76 46.8 25.1 37.7#





#15
Acrylonitrile
Concen: 211.303 ug/l
RT: 5.161 min Scan# 548
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 9605
Ion Ratio Lower Upper
53 100
52 81.2 67.0 100.4
51 35.5 29.8 44.6


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#16
Acetone
Concen: 218.501 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 43 Resp: 80418
Ion Ratio Lower Upper
43 100
58 32.0 22.0 33.0





#17
Carbon Disulfide
Concen: 25.456 ug/l
RT: 4.198 min Scan# 390
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 11172
Ion Ratio Lower Upper
76 100
78 8.7 7.4 11.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#18
Methyl Acetate
Concen: 37.408 ug/l
RT: 4.472 min Scan# 435
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 43 Resp: 38611
Ion Ratio Lower Upper
43 100
74 29.2 16.6 24.8#





#19
Methyl tert-butyl Ether
Concen: 42.326 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :



VSTDCCC050

Tgt Ion: 73 Resp: 194420
Ion Ratio Lower Upper
73 100
57 19.9 18.6 27.8

Manual Integrations



APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#20
Methylene Chloride
Concen: 35.929 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 84 Resp: 72096
Ion Ratio Lower Upper
84 100
49 109.8 114.4 171.6#
51 35.3 35.3 52.9
86 67.4 53.7 80.5







#21
trans-1,2-Dichloroethene
Concen: 35.740 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 96 Resp: 63290
Ion Ratio Lower Upper
96 100
61 120.8 120.5 180.7
98 63.1 49.8 74.8



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#22
Diisopropyl ether
Concen: 42.428 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 45 Resp: 194145
Ion Ratio Lower Upper
45 100
43 59.8 49.3 73.9
87 34.3 19.2 28.8#
59 12.6 8.4 12.6#





#23

Vinyl Acetate

Concen: 209.543 ug/l

RT: 6.058 min Scan# 695

Delta R.T. -0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 65737

Ion Ratio Lower Upper

43 100



86 12.9 7.1 10.7

Manual Integrations

APPROVED



Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#24

1,1-Dichloroethane

Concen: 41.653 ug/l

RT: 6.021 min Scan# 689

Delta R.T. -0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40




Tgt Ion: 63 Resp: 118640


Ion Ratio Lower Upper

63 100

98 7.4 3.6 10.8

100 5.1 2.0 5.8







#25
2-Butanone
Concen: 212.262 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 120860
Ion Ratio Lower Upper
43 100
72 32.3 17.7 26.5



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#26
2,2-Dichloropropane
Concen: 43.931 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 77 Resp: 115894
Ion Ratio Lower Upper
77 100
97 23.7 10.4 31.4





#27

cis-1,2-Dichloroethene

Concen: 41.346 ug/l

RT: 6.990 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 83750



Ion	Ratio	Lower	Upper
96	100		
61	127.8	0.0	319.8
98	64.6	0.0	127.0

Manual Integrations

APPROVED



Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#28

Bromochloromethane

Concen: 45.094 ug/l

RT: 7.332 min Scan# 904




Delta R.T. -0.006 min


Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Tgt Ion: 49 Resp: 51810


Ion	Ratio	Lower	Upper
49	100		
129	2.5	0.0	2.2#
130	91.1	55.4	83.0#





#29
Tetrahydrofuran
Concen: 207.677 ug/l
RT: 7.344 min Scan# 906
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 42 Resp: 76162
Ion Ratio Lower Upper
42 100
72 54.6 30.3 45.5
71 49.8 29.0 43.4


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#30
Chloroform
Concen: 43.167 ug/l
RT: 7.508 min Scan# 933
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 83 Resp: 135079
Ion Ratio Lower Upper
83 100
85 66.2 51.8 77.6





#31
Cyclohexane
Concen: 33.236 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 8633
Ion Ratio Lower Upper
56 100
69 36.3 26.7 40.1
84 100.5 67.2 100.8


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#32
1,1,1-Trichloroethane
Concen: 42.737 ug/l
RT: 7.704 min Scan# 965
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Tgt Ion: 97 Resp: 117100
Ion Ratio Lower Upper
97 100
99 64.8 50.7 76.1
61 40.8 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 36.590 ug/l

RT: 8.143 min Scan# 1037

Delta R.T. 0.001 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 5781

Ion Ratio Lower Upper

65 100



67 53.9 0.0 99.4

Manual Integrations

APPROVED



Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40




Tgt Ion:114 Resp: 226529


Ion Ratio Lower Upper

114 100

63 16.1 0.0 43.2


88 14.6 0.0 31.8





#35
Dibromofluoromethane
Concen: 42.001 ug/l
RT: 7.716 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:113 Resp: 61404
Ion Ratio Lower Upper
113 100
111 103.4 82.6 124.0
192 18.4 17.6 26.4


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#36
1,1-Dichloropropene
Concen: 38.478 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 75 Resp: 89482
Ion Ratio Lower Upper
75 100
110 37.5 17.1 51.2
77 31.5 24.4 36.6





#37
Ethyl Acetate
Concen: 41.429 ug/l
RT: 7.076 min Scan# 861
Delta R.T. 0.007 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 5328
Ion Ratio Lower Upper
43 100
61 14.9 9.7 14.5
70 12.1 7.1 10.7


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#38
Carbon Tetrachloride
Concen: 42.164 ug/l
RT: 7.899 min Scan# 997
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:117 Resp: 101876
Ion Ratio Lower Upper
117 100
119 97.0 77.9 116.9
121 31.7 24.8 37.2





#39

Methylcyclohexane

Concen: 36.842 ug/l

RT: 9.185 min Scan# 1208

Delta R.T. -0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 10970

Ion Ratio Lower Upper

83 100

55 63.2 66.5 99.7



98 51.6 37.9 56.9

Manual Integrations

APPROVED



Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#40

Benzene

Concen: 40.288 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY008434.D




Acq: 21 Apr 2022 10:40


Tgt Ion: 78 Resp: 273339

Ion Ratio Lower Upper

78 100


77 24.2 20.4 30.6





#41
Methacrylonitrile
Concen: 42.306 ug/l m
RT: 7.313 min Scan# 901
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 26380
Ion Ratio Lower Upper
41 100
39 137.2 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#42
1,2-Dichloroethane
Concen: 41.781 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 62 Resp: 85344
Ion Ratio Lower Upper
62 100
98 10.7 0.0 14.8





#43
Isopropyl Acetate
Concen: 43.824 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 103658
Ion Ratio Lower Upper
43 100
61 31.5 20.2 30.2
87 16.8 8.9 13.3


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#44
Trichloroethene
Concen: 41.867 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:130 Resp: 78256
Ion Ratio Lower Upper
130 100
95 97.0 0.0 188.2





#45

1,2-Dichloropropane

Concen: 44.390 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 7232

Ion Ratio Lower Upper

63 100



65 30.7 23.0 34.6

Manual Integrations

APPROVED



Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#46

Dibromomethane

Concen: 41.975 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY008434.D

Acq: 21 Apr 2022 10:40




Tgt Ion: 93 Resp: 47912


Ion Ratio Lower Upper

93 100

95 81.0 64.3 96.5


174 94.7 83.4 125.0





#47
Bromodichloromethane
Concen: 45.006 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 10905
Ion Ratio Lower Upper
83 100
85 65.1 51.9 77.9
127 8.8 6.6 10.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#48
Methyl methacrylate
Concen: 43.253 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 41 Resp: 43501
Ion Ratio Lower Upper
41 100
69 108.7 58.2 87.2#
39 65.2 47.4 71.0





#49
1,4-Dioxane
Concen: 864.691 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 88 Resp: 12440
Ion Ratio Lower Upper
88 100
43 26.2 32.8 49.2
58 59.1 59.8 89.6


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#50
Toluene-d8
Concen: 36.547 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 98 Resp: 209381
Ion Ratio Lower Upper
98 100
100 64.7 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 225.483 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 280569
Ion Ratio Lower Upper
43 100
58 43.9 28.1 42.1


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#52
Toluene
Concen: 41.838 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 92 Resp: 187719
Ion Ratio Lower Upper
92 100
91 170.5 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 44.129 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :



VSTDCCC050

Tgt Ion: 75 Resp: 114669
Ion Ratio Lower Upper
75 100
77 32.1 25.4 38.2

Manual Integrations



APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#54
cis-1,3-Dichloropropene
Concen: 42.758 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 75 Resp: 123122
Ion Ratio Lower Upper
75 100
77 30.9 25.6 38.4
39 39.7 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 44.488 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 97 Resp: 69308
Ion Ratio Lower Upper
97 100
83 82.5 69.8 104.8
85 53.8 44.3 66.5
99 63.2 49.8 74.6


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#56
Ethyl methacrylate
Concen: 44.935 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 69 Resp: 90390
Ion Ratio Lower Upper
69 100
41 53.7 63.4 95.0#
39 32.8 35.8 53.8#





#57
1,3-Dichloropropane
Concen: 43.713 ug/l
RT: 10.794 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050



Tgt Ion: 76 Resp: 111169
Ion Ratio Lower Upper
76 100
78 32.7 25.0 37.6


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin Yesilyurt


04/22/2022



#58
2-Chloroethyl Vinyl ether
Concen: 138.628 ug/l
RT: 9.782 min Scan# 1306
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 63 Resp: 152829
Ion Ratio Lower Upper
63 100
106 32.1 21.5 32.3







#59
2-Hexanone
Concen: 229.694 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 196680
Ion Ratio Lower Upper
43 100
58 62.3 24.4 73.4



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#60
Dibromochloromethane
Concen: 46.504 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:129 Resp: 83237
Ion Ratio Lower Upper
129 100
127 77.1 38.7 116.1





#61
1,2-Dibromoethane
Concen: 42.643 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :

MSVOA_Y

ClientSampleId :



VSTDCCC050

Tgt Ion:107 Resp: 6570
Ion Ratio Lower Upper
107 100
109 95.9 75.4 113.2

Manual Integrations



APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#62
4-Bromofluorobenzene
Concen: 39.916 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion: 95 Resp: 78579
Ion Ratio Lower Upper
95 100
174 83.4 0.0 177.0
176 81.0 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 20808
Ion Ratio Lower Upper
117 100
82 51.7 43.0 64.4
119 30.9 25.4 38.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#64
Tetrachloroethene
Concen: 41.469 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion:164 Resp: 63993
Ion Ratio Lower Upper
164 100
166 123.4 100.4 150.6
129 96.6 71.4 107.0
131 89.8 70.8 106.2





#65
Chlorobenzene
Concen: 45.167 ug/l
RT: 11.514 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 212514
Ion Ratio Lower Upper
112 100
114 32.0 25.8 38.6


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 48.103 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion:131 Resp: 83561
Ion Ratio Lower Upper
131 100
133 97.1 47.9 143.6
119 67.2 30.8 92.3





#67
Ethyl Benzene
Concen: 45.593 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 37230
Ion Ratio Lower Upper
91 100
106 31.9 24.5 36.7


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#68
m/p-Xylenes
Concen: 88.564 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:106 Resp: 289679
Ion Ratio Lower Upper
106 100
91 196.6 166.6 249.8







#69
o-Xylene
Concen: 44.726 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 14084
Ion Ratio Lower Upper
106 100
91 205.6 110.3 331.0



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#70
Styrene
Concen: 45.642 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion:104 Resp: 242205
Ion Ratio Lower Upper
104 100
78 45.7 42.4 63.6
103 54.3 44.4 66.6





#71
Bromoform
Concen: 47.815 ug/l
RT: 12.209 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:173 Resp: 5205
Ion Ratio Lower Upper
173 100
175 47.6 24.6 74.0
254 0.0 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion:152 Resp: 101505
Ion Ratio Lower Upper
152 100
115 58.1 28.2 84.7
150 175.6 0.0 347.6





#73
Isopropylbenzene
Concen: 48.805 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 375602
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.2


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#74
N-ethyl acetate
Concen: 49.331 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40




Tgt Ion: 43 Resp: 93807
Ion Ratio Lower Upper
43 100
70 50.2 28.7 43.1#
55 24.9 17.3 25.9
61 27.7 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 50.536 ug/l
RT: 12.587 min Scan# 1766
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 8689
Ion Ratio Lower Upper
83 100
131 10.4 5.6 16.8
85 64.3 32.3 96.8


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#76
1,2,3-Trichloropropane
Concen: 40.570 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.007 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 75 Resp: 53294
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0







#77
Bromobenzene
Concen: 46.903 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:156 Resp: 8803
Ion Ratio Lower Upper
156 100
77 180.9 95.5 286.4
158 96.4 48.0 144.0



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#78
n-propylbenzene
Concen: 47.993 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 91 Resp: 443034
Ion Ratio Lower Upper
91 100
120 23.3 11.3 33.9







#79
2-Chlorotoluene
Concen: 47.710 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 247878
Ion Ratio Lower Upper
91 100
126 35.4 16.9 50.6



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#80
1,3,5-Trimethylbenzene
Concen: 48.141 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion:105 Resp: 307531
Ion Ratio Lower Upper
105 100
120 49.4 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 50.707 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 2871
Ion Ratio Lower Upper
75 100
53 81.5 85.3 127.9
89 49.5 37.0 55.4


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#82
4-Chlorotoluene
Concen: 47.598 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion: 91 Resp: 256263
Ion Ratio Lower Upper
91 100
126 35.7 17.0 51.0







#83
tert-Butylbenzene
Concen: 50.027 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:119 Resp: 27742
Ion Ratio Lower Upper
119 100
91 62.8 34.8 104.4
134 27.5 13.3 39.9



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#84
1,2,4-Trimethylbenzene
Concen: 48.885 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion:105 Resp: 309848
Ion Ratio Lower Upper
105 100
120 46.0 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 49.937 ug/l
 RT: 13.257 min Scan# 1876
 Delta R.T. 0.006 min
 Lab File: VY008434.D
 Acq: 21 Apr 2022 10:40



Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion:105 Resp: 41792
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.2 30.4


Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
 Supervised By :Semsettin
 Yesilyurt



04/22/2022




#86
 p-Isopropyltoluene
 Concen: 50.185 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY008434.D
 Acq: 21 Apr 2022 10:40



Tgt Ion:119 Resp: 351739
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.2 39.5
 91 22.3 12.4 37.2







#87
1,3-Dichlorobenzene
Concen: 47.532 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 17861
Ion Ratio Lower Upper
146 100
111 39.3 19.5 58.5
148 62.9 31.6 95.0



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#88
1,4-Dichlorobenzene
Concen: 47.478 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Tgt Ion:146 Resp: 176731
Ion Ratio Lower Upper
146 100
111 38.0 19.3 57.8
148 63.3 32.6 97.8







#89
n-Butylbenzene
Concen: 49.877 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 31728
Ion Ratio Lower Upper
91 100
92 53.1 26.3 78.9
134 26.5 13.0 38.9



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#90
Hexachloroethane
Concen: 53.264 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion:117 Resp: 67117
Ion Ratio Lower Upper
117 100
201 68.3 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 47.820 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 159779
Ion Ratio Lower Upper
146 100
111 40.3 20.3 60.9
148 63.2 32.3 96.9


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt



04/22/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 47.597 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Tgt Ion: 75 Resp: 12300
Ion Ratio Lower Upper
75 100
155 96.6 38.0 113.9
157 122.6 47.8 143.3







#93
1,2,4-Trichlorobenzene
Concen: 49.209 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:180 Resp: 9152
Ion Ratio Lower Upper
180 100
182 93.5 47.3 141.9
145 30.1 14.6 43.8



Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022






04/22/2022
Supervised By :Semsettin
Yesilyurt


04/22/2022



#94
Hexachlorobutadiene
Concen: 50.847 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40


Tgt Ion:225 Resp: 46142
Ion Ratio Lower Upper
225 100
223 62.4 30.8 92.3
227 62.2 31.6 94.7





#95
Naphthalene
Concen: 48.527 ug/l
RT: 15.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 20460
Ion Ratio Lower Upper
128 100
127 13.1 11.0 16.4
129 11.0 8.4 12.6


Manual Integrations APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




04/22/2022
Supervised By :Semsettin
Yesilyurt

04/22/2022



#96
1,2,3-Trichlorobenzene
Concen: 48.926 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008434.D
Acq: 21 Apr 2022 10:40



Tgt Ion:180 Resp: 79646
Ion Ratio Lower Upper
180 100
182 94.2 47.3 141.9
145 31.7 15.3 45.9

