

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008443.D
 Acq On : 21 Apr 2022 15:59
 Operator : KP/MD
 Sample : VY0421SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0421SBS01

Quant Time: Apr 22 05:25:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	85108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	150355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	145745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	53933	53.858	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 107.720%			
35) Dibromofluoromethane	7.716	113	48628	50.113	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.220%			
50) Toluene-d8	10.179	98	169350	44.062	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 88.120%			
62) 4-Bromofluorobenzene	12.483	95	60299	46.148	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 92.300%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	15874	20.539	ug/l	99
3) Chloromethane	2.113	50	28294	23.265	ug/l	98
4) Vinyl Chloride	2.253	62	33951	21.485	ug/l	97
5) Bromomethane	2.643	94	33184	22.584	ug/l	99
6) Chloroethane	2.796	64	23319	22.181	ug/l	96
7) Trichlorofluoromethane	3.125	101	35332	21.583	ug/l	93
8) Diethyl Ether	3.533	74	13459	23.888	ug/l	72
9) 1,1,2-Trichlorotrifluo...	3.899	101	21404	21.964	ug/l	93
10) Methyl Iodide	4.094	142	22644	18.732	ug/l	90
11) Tert butyl alcohol	4.948	59	15044	137.331	ug/l #	91
12) 1,1-Dichloroethene	3.875	96	19752	21.469	ug/l	81
13) Acrolein	3.729	56	13967	131.003	ug/l	95
14) Allyl chloride	4.478	41	25437	21.183	ug/l #	85
15) Acrylonitrile	5.161	53	34252	118.898	ug/l	99
16) Acetone	3.948	43	26875	115.222	ug/l	91
17) Carbon Disulfide	4.192	76	57623	20.717	ug/l	100
18) Methyl Acetate	4.478	43	15286	23.369	ug/l #	86
19) Methyl tert-butyl Ether	5.216	73	65807	22.606	ug/l	96
20) Methylene Chloride	4.716	84	32053	23.104	ug/l #	78
21) trans-1,2-Dichloroethene	5.216	96	24420	21.760	ug/l	83
22) Diisopropyl ether	6.118	45	64105	22.106	ug/l #	94
23) Vinyl Acetate	6.064	43	231549	116.464	ug/l #	88
24) 1,1-Dichloroethane	6.015	63	38946	21.575	ug/l	98
25) 2-Butanone	6.984	43	42564	117.956	ug/l #	80
26) 2,2-Dichloropropane	6.978	77	35936	21.495	ug/l	94
27) cis-1,2-Dichloroethene	6.984	96	28728	22.379	ug/l	81
28) Bromochloromethane	7.338	49	16133	22.157	ug/l #	74
29) Tetrahydrofuran	7.344	42	27804	119.631	ug/l #	75
30) Chloroform	7.508	83	43358	21.864	ug/l	99
31) Cyclohexane	7.783	56	34763	21.116	ug/l #	82
32) 1,1,1-Trichloroethane	7.703	97	37135	21.385	ug/l	96
36) 1,1-Dichloropropene	7.923	75	32569	21.100	ug/l	97
37) Ethyl Acetate	7.069	43	20783	24.346	ug/l #	94
38) Carbon Tetrachloride	7.905	117	32666	20.369	ug/l	95
39) Methylcyclohexane	9.185	83	41287	20.891	ug/l	86
40) Benzene	8.161	78	94403	20.964	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
 Data File : VY008443.D
 Acq On : 21 Apr 2022 15:59
 Operator : KP/MD
 Sample : VY0421SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0421SBS01

Quant Time: Apr 22 05:25:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	9421m	22.758	ug/l	
42) 1,2-Dichloroethane	8.234	62	30373	22.403	ug/l	90
43) Isopropyl Acetate	8.277	43	35816	22.813	ug/l #	86
44) Trichloroethene	8.941	130	26582	21.426	ug/l	98
45) 1,2-Dichloropropane	9.215	63	23775	21.986	ug/l	99
46) Dibromomethane	9.307	93	16525	21.812	ug/l	94
47) Bromodichloromethane	9.496	83	34775	21.623	ug/l	98
48) Methyl methacrylate	9.295	41	15401	23.072	ug/l #	77
49) 1,4-Dioxane	9.295	88	4042	423.091	ug/l #	90
51) 4-Methyl-2-Pentanone	10.069	43	97362	117.888	ug/l #	87
52) Toluene	10.246	92	62112	20.857	ug/l	97
53) t-1,3-Dichloropropene	10.465	75	37872	21.959	ug/l	96
54) cis-1,3-Dichloropropene	9.929	75	41688	21.812	ug/l #	84
55) 1,1,2-Trichloroethane	10.642	97	22627	21.882	ug/l	95
56) Ethyl methacrylate	10.514	69	28959	21.690	ug/l #	80
57) 1,3-Dichloropropane	10.794	76	38007	22.516	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	80662	110.235	ug/l	90
59) 2-Hexanone	10.831	43	66955	117.809	ug/l	84
60) Dibromochloromethane	10.983	129	26031	21.911	ug/l	98
61) 1,2-Dibromoethane	11.087	107	22974	22.464	ug/l	99
64) Tetrachloroethene	10.721	164	22577	20.889	ug/l	99
65) Chlorobenzene	11.520	112	68706	20.849	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.587	131	24963	20.517	ug/l	98
67) Ethyl Benzene	11.593	91	116363	20.346	ug/l	97
68) m/p-Xylenes	11.703	106	91860	40.098	ug/l	92
69) o-Xylene	12.032	106	44351	20.109	ug/l	93
70) Styrene	12.044	104	74071	19.929	ug/l	94
71) Bromoform	12.209	173	16162	21.197	ug/l #	98
73) Isopropylbenzene	12.331	105	113485	20.862	ug/l	99
74) N-amyl acetate	12.148	43	30429	22.639	ug/l #	81
75) 1,1,2,2-Tetrachloroethane	12.581	83	27641	22.741	ug/l	98
76) 1,2,3-Trichloropropane	12.636	75	17260m	18.588	ug/l	
77) Bromobenzene	12.611	156	28292	21.325	ug/l	95
78) n-propylbenzene	12.672	91	137483	21.070	ug/l	99
79) 2-Chlorotoluene	12.757	91	76605	20.859	ug/l	97
80) 1,3,5-Trimethylbenzene	12.812	105	94827	21.001	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.379	75	9197	22.978	ug/l	86
82) 4-Chlorotoluene	12.855	91	78616	20.658	ug/l	95
83) tert-Butylbenzene	13.074	119	82218	20.975	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	94243	21.036	ug/l	99
85) sec-Butylbenzene	13.251	105	123180	20.823	ug/l	100
86) p-Isopropyltoluene	13.373	119	102851	20.761	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	54992	20.704	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	54838	20.842	ug/l	99
89) n-Butylbenzene	13.696	91	95625	21.267	ug/l	98
90) Hexachloroethane	13.965	117	18619	20.904	ug/l	82
91) 1,2-Dichlorobenzene	13.739	146	49954	21.151	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.361	75	4068	22.271	ug/l	81
93) 1,2,4-Trichlorobenzene	15.007	180	28428	21.624	ug/l	97
94) Hexachlorobutadiene	15.111	225	14071	21.937	ug/l	98
95) Naphthalene	15.239	128	67184	22.543	ug/l	98
96) 1,2,3-Trichlorobenzene	15.428	180	25985	22.583	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042122\
Data File : VY008443.D
Acq On : 21 Apr 2022 15:59
Operator : KP/MD
Sample : VY0421SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022

Quant Time: Apr 22 05:25:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion:168 Resp: 85108
Ion Ratio Lower Upper
168 100
99 54.5 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#2
Dichlorodifluoromethane
Concen: 20.539 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 85 Resp: 15874
Ion Ratio Lower Upper
85 100
87 31.6 16.1 48.2





#3
Chloromethane
Concen: 23.265 ug/l
RT: 2.113 min Scan# 48
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion: 50 Resp: 28294
Ion Ratio Lower Upper
50 100
52 31.7 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#4
Vinyl Chloride
Concen: 21.485 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 62 Resp: 33951
Ion Ratio Lower Upper
62 100
64 33.1 25.0 37.6





#5
Bromomethane
Concen: 22.584 ug/l
RT: 2.643 min Scan# 11
Delta R.T. -0.007 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 94 Resp: 33184
Ion Ratio Lower Upper
94 100
96 91.4 73.5 110.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#6
Chloroethane
Concen: 22.181 ug/l
RT: 2.796 min Scan# 160
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 64 Resp: 23319
Ion Ratio Lower Upper
64 100
66 29.5 25.4 38.0




Abundance Scan 214 (3.125 min): VY008438.D\data.ms (-20)




m/z-->

Abundance Scan 214 (3.125 min): VY008443.D\data.ms




m/z-->

Abundance Scan 214 (3.125 min): VY008443.D\data.ms (-16)




m/z-->

Abundance Scan 280 (3.527 min): VY008438.D\data.ms (-27)




m/z-->

Abundance Scan 281 (3.533 min): VY008443.D\data.ms



m/z-->

Abundance Scan 281 (3.533 min): VY008443.D\data.ms (-23)



m/z-->

#7

Trichlorofluoromethane

Concen: 21.583 ug/l

RT: 3.125 min Scan# 214

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Tgt Ion: 101 Resp: 3533

Ion Ratio Lower Upper

101 100

103 65.0 48.0 72.0

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01


Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022

Abundance



Time-->

#8

Diethyl Ether

Concen: 23.888 ug/l

RT: 3.533 min Scan# 281

Delta R.T. 0.006 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59


Tgt Ion: 74 Resp: 13459

Ion Ratio Lower Upper


74 100

45 80.8 55.1 165.5

Abundance



Time-->







#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.964 ug/l
RT: 3.899 min Scan# 341
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion:101 Resp: 21404
Ion Ratio Lower Upper
101 100
85 43.9 40.6 60.8
151 75.9 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#10
Methyl Iodide
Concen: 18.732 ug/l
RT: 4.094 min Scan# 373
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion:142 Resp: 22644
Ion Ratio Lower Upper
142 100
127 43.1 41.5 62.3
141 14.1 11.9 17.9









#11
Tert butyl alcohol
Concen: 137.331 ug/l
RT: 4.948 min Scan# 513
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 59 Resp: 15044
Ion Ratio Lower Upper
59 100
57 5.0 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#12
1,1-Dichloroethene
Concen: 21.469 ug/l
RT: 3.875 min Scan# 337
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 96 Resp: 19752
Ion Ratio Lower Upper
96 100
61 146.7 145.8 218.6
98 68.9 52.5 78.7





#13
Acrolein
Concen: 131.003 ug/l
RT: 3.729 min Scan# 313
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 56 Resp: 1396
Ion Ratio Lower Upper
56 100
55 67.8 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#14
Allyl chloride
Concen: 21.183 ug/l
RT: 4.478 min Scan# 436
Delta R.T. -0.001 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 41 Resp: 25437
Ion Ratio Lower Upper
41 100
39 79.9 58.8 88.2
76 48.5 25.1 37.7#





#15
Acrylonitrile
Concen: 118.898 ug/l
RT: 5.161 min Scan# 548
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 53 Resp: 3425
Ion Ratio Lower Upper
53 100
52 82.4 67.0 100.4
51 36.5 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#16
Acetone
Concen: 115.222 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 43 Resp: 26875
Ion Ratio Lower Upper
43 100
58 32.1 22.0 33.0





#17
Carbon Disulfide
Concen: 20.717 ug/l
RT: 4.192 min Scan# 389
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 76 Resp: 5762
Ion Ratio Lower Upper
76 100
78 9.3 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#18
Methyl Acetate
Concen: 23.369 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 43 Resp: 15286
Ion Ratio Lower Upper
43 100
74 27.2 16.6 24.8





#19

Methyl tert-butyl Ether

Concen: 22.606 ug/l

RT: 5.216 min Scan# 51

Delta R.T. -0.006 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 73 Resp: 6580

Ion Ratio Lower Upper

73 100





57 21.1 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#20

Methylene Chloride

Concen: 23.104 ug/l

RT: 4.716 min Scan# 475

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Tgt Ion: 84 Resp: 32053




Ion Ratio Lower Upper


84 100

49 104.2 114.4 171.6#

51 33.7 35.3 52.9#

86 62.7 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 21.760 ug/l
RT: 5.216 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 96 Resp: 24420
Ion Ratio Lower Upper
96 100
61 121.3 120.5 180.7
98 59.4 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#22
Diisopropyl ether
Concen: 22.106 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 45 Resp: 64105
Ion Ratio Lower Upper
45 100
43 59.9 49.3 73.9
87 31.1 19.2 28.8#
59 12.8 8.4 12.6#





#23
 Vinyl Acetate
 Concen: 116.464 ug/l
 RT: 6.064 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: VY008443.D
 Acq: 21 Apr 2022 15:59




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0421SBS01



Tgt Ion: 43 Resp: 231549
 Ion Ratio Lower Upper
 43 100
 86 13.3 7.1 10.7




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022



#24
 1,1-Dichloroethane
 Concen: 21.575 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. -0.006 min
 Lab File: VY008443.D
 Acq: 21 Apr 2022 15:59

Tgt Ion: 63 Resp: 38946
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.6 10.8
 100 5.4 2.0 5.8





#25

2-Butanone

Concen: 117.956 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 43 Resp: 42564

Ion Ratio Lower Upper

43 100





72 31.6 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#26

2,2-Dichloropropane

Concen: 21.495 ug/l

RT: 6.978 min Scan# 846

Delta R.T. -0.000 min

Lab File: VY008443.D




Acq: 21 Apr 2022 15:59


Tgt Ion: 77 Resp: 35936

Ion Ratio Lower Upper

77 100

97 23.7 10.4 31.4









#27
cis-1,2-Dichloroethene
Concen: 22.379 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 96 Resp: 28728
Ion Ratio Lower Upper
96 100
61 125.2 0.0 319.8
98 65.1 0.0 127.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#28
Bromochloromethane
Concen: 22.157 ug/l
RT: 7.338 min Scan# 905
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 49 Resp: 16133
Ion Ratio Lower Upper
49 100
129 2.2 0.0 2.2#
130 91.0 55.4 83.0#





#29
Tetrahydrofuran
Concen: 119.631 ug/l
RT: 7.344 min Scan# 906
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 42 Resp: 27804
Ion Ratio Lower Upper
42 100
72 52.9 30.3 45.5
71 51.1 29.0 43.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#30
Chloroform
Concen: 21.864 ug/l
RT: 7.508 min Scan# 933
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 83 Resp: 43358
Ion Ratio Lower Upper
83 100
85 65.8 51.8 77.6





#31
Cyclohexane
Concen: 21.116 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 56 Resp: 3476
Ion Ratio Lower Upper
56 100
69 38.7 26.7 40.1
84 103.7 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#32
1,1,1-Trichloroethane
Concen: 21.385 ug/l
RT: 7.703 min Scan# 965
Delta R.T. -0.001 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 97 Resp: 37135
Ion Ratio Lower Upper
97 100
99 65.4 50.7 76.1
61 42.5 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 53.858 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 65 Resp: 5393

Ion Ratio Lower Upper

65 100





67 52.7 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59




Tgt Ion:114 Resp: 150355


Ion Ratio Lower Upper

114 100

63 18.4 0.0 43.2


88 14.6 0.0 31.8





#35
Dibromofluoromethane
Concen: 50.113 ug/l
RT: 7.716 min Scan# 901
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:113 Resp: 48628
Ion Ratio Lower Upper
113 100
111 103.8 82.6 124.0
192 17.7 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#36
1,1-Dichloropropene
Concen: 21.100 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 75 Resp: 32569
Ion Ratio Lower Upper
75 100
110 36.0 17.1 51.2
77 32.3 24.4 36.6





#37
Ethyl Acetate
Concen: 24.346 ug/l
RT: 7.069 min Scan# 861
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion: 43 Resp: 2078
Ion Ratio Lower Upper
43 100
61 14.7 9.7 14.5
70 10.9 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#38
Carbon Tetrachloride
Concen: 20.369 ug/l
RT: 7.905 min Scan# 998
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion:117 Resp: 32666
Ion Ratio Lower Upper
117 100
119 90.8 77.9 116.9
121 30.7 24.8 37.2





#39

Methylcyclohexane

Concen: 20.891 ug/l

RT: 9.185 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 83 Resp: 4128

Ion Ratio Lower Upper

83 100

55 68.8 66.5 99.7





98 53.6 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#40

Benzene

Concen: 20.964 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY008443.D




Acq: 21 Apr 2022 15:59


Tgt Ion: 78 Resp: 94403

Ion Ratio Lower Upper

78 100


77 23.7 20.4 30.6





#41
Methacrylonitrile
Concen: 22.758 ug/l m
RT: 7.319 min Scan# 902
Delta R.T. 0.012 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion: 41 Resp: 942
Ion Ratio Lower Upper
41 100
39 140.2 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#42
1,2-Dichloroethane
Concen: 22.403 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 62 Resp: 30373
Ion Ratio Lower Upper
62 100
98 10.8 0.0 14.8





#43
Isopropyl Acetate
Concen: 22.813 ug/l
RT: 8.277 min Scan# 1059
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion: 43 Resp: 35810
Ion Ratio Lower Upper
43 100
61 32.0 20.2 30.2
87 16.4 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#44
Trichloroethene
Concen: 21.426 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion:130 Resp: 26582
Ion Ratio Lower Upper
130 100
95 96.4 0.0 188.2





#45

1,2-Dichloropropane

Concen: 21.986 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 63 Resp: 2377

Ion Ratio Lower Upper

63 100





65 29.5 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#46

Dibromomethane

Concen: 21.812 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59




Tgt Ion: 93 Resp: 16525


Ion Ratio Lower Upper

93 100

95 85.3 64.3 96.5

174 96.9 83.4 125.0









#47
Bromodichloromethane
Concen: 21.623 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 83 Resp: 34775
Ion Ratio Lower Upper
83 100
85 63.0 51.9 77.9
127 7.8 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#48
Methyl methacrylate
Concen: 23.072 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 41 Resp: 15401
Ion Ratio Lower Upper
41 100
69 106.6 58.2 87.2#
39 61.2 47.4 71.0





#49
1,4-Dioxane
Concen: 423.091 ug/l
RT: 9.295 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 88 Resp: 404
Ion Ratio Lower Upper
88 100
43 30.4 32.8 49.2
58 69.6 59.8 89.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#50
Toluene-d8
Concen: 44.062 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 98 Resp: 169350
Ion Ratio Lower Upper
98 100
100 66.0 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 117.888 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 43 Resp: 97362
Ion Ratio Lower Upper
43 100
58 42.4 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#52
Toluene
Concen: 20.857 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 92 Resp: 62112
Ion Ratio Lower Upper
92 100
91 173.6 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 21.959 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 75 Resp: 37872
Ion Ratio Lower Upper
75 100
77 33.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#54
cis-1,3-Dichloropropene
Concen: 21.812 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 75 Resp: 41688
Ion Ratio Lower Upper
75 100
77 32.1 25.6 38.4
39 40.5 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 21.882 ug/l
RT: 10.642 min Scan# 1448
Delta R.T. -0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 97 Resp: 2262
Ion Ratio Lower Upper
97 100
83 81.4 69.8 104.8
85 50.6 44.3 66.5
99 61.2 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#56
Ethyl methacrylate
Concen: 21.690 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 69 Resp: 28959
Ion Ratio Lower Upper
69 100
41 58.7 63.4 95.0#
39 35.2 35.8 53.8#





#57

1,3-Dichloropropane

Concen: 22.516 ug/l

RT: 10.794 min Scan# 1471

Delta R.T. 0.006 min

Lab File: VY008443.D

Acq: 21 Apr 2022 15:59

Instrument :

MSVOA_Y

ClientSampleId :

VY0421SBS01

Tgt Ion: 76 Resp: 3800

Ion Ratio Lower Upper

76 100





78 32.6 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/22/2022

Supervised By :Semsettin Yesilyurt 04/22/2022



#58

2-Chloroethyl Vinyl ether

Concen: 110.235 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.000 min

Lab File: VY008443.D




Acq: 21 Apr 2022 15:59


Tgt Ion: 63 Resp: 80662

Ion Ratio Lower Upper

63 100

106 31.8 21.5 32.3









#59
2-Hexanone
Concen: 117.809 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 43 Resp: 6695
Ion Ratio Lower Upper
43 100
58 59.8 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#60
Dibromochloromethane
Concen: 21.911 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion:129 Resp: 26031
Ion Ratio Lower Upper
129 100
127 75.9 38.7 116.1









#61
1,2-Dibromoethane
Concen: 22.464 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion:107 Resp: 22974
Ion Ratio Lower Upper
107 100
109 93.2 75.4 113.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#62
4-Bromofluorobenzene
Concen: 46.148 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 95 Resp: 60299
Ion Ratio Lower Upper
95 100
174 85.5 0.0 177.0
176 80.8 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:117 Resp: 14574
Ion Ratio Lower Upper
117 100
82 51.6 43.0 64.4
119 31.0 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#64
Tetrachloroethene
Concen: 20.889 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion:164 Resp: 22577
Ion Ratio Lower Upper
164 100
166 127.6 100.4 150.6
129 90.1 71.4 107.0
131 88.1 70.8 106.2





#65
Chlorobenzene
Concen: 20.849 ug/l
RT: 11.520 min Scan# 11520
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:112 Resp: 68700
Ion Ratio Lower Upper
112 100
114 30.6 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 20.517 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion:131 Resp: 24963
Ion Ratio Lower Upper
131 100
133 96.0 47.9 143.6
119 65.7 30.8 92.3





#67
Ethyl Benzene
Concen: 20.346 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 91 Resp: 11636
Ion Ratio Lower Upper
91 100
106 32.2 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#68
m/p-Xylenes
Concen: 40.098 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion:106 Resp: 91860
Ion Ratio Lower Upper
106 100
91 195.8 166.6 249.8









#69
o-Xylene
Concen: 20.109 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion:106 Resp: 4435
Ion Ratio Lower Upper
106 100
91 209.3 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#70
Styrene
Concen: 19.929 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion:104 Resp: 74071
Ion Ratio Lower Upper
104 100
78 46.1 42.4 63.6
103 54.1 44.4 66.6





#71
Bromoform
Concen: 21.197 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:173 Resp: 16162
Ion Ratio Lower Upper
173 100
175 48.2 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion:152 Resp: 71748
Ion Ratio Lower Upper
152 100
115 57.5 28.2 84.7
150 164.3 0.0 347.6





#73
Isopropylbenzene
Concen: 20.862 ug/l
RT: 12.331 min Scan# 11348
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:105 Resp: 11348
Ion Ratio Lower Upper
105 100
120 27.2 13.4 40.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#74
N-ethyl acetate
Concen: 22.639 ug/l
RT: 12.148 min Scan# 1694
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 43 Resp: 30429
Ion Ratio Lower Upper
43 100
70 51.3 28.7 43.1#
55 26.2 17.3 25.9#
61 28.2 17.0 25.6#









#75
1,1,2,2-Tetrachloroethane
Concen: 22.741 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 83 Resp: 2764
Ion Ratio Lower Upper
83 100
131 9.8 5.6 16.8
85 63.4 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#76
1,2,3-Trichloropropane
Concen: 18.588 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.007 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion: 75 Resp: 17260
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 21.325 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion: 156 Resp: 28292
Ion Ratio Lower Upper
156 100
77 181.5 95.5 286.4
158 97.9 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#78
n-propylbenzene
Concen: 21.070 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion: 91 Resp: 137483
Ion Ratio Lower Upper
91 100
120 23.0 11.3 33.9









#79
2-Chlorotoluene
Concen: 20.859 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. -0.001 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 91 Resp: 7660
Ion Ratio Lower Upper
91 100
126 35.6 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#80
1,3,5-Trimethylbenzene
Concen: 21.001 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion:105 Resp: 94827
Ion Ratio Lower Upper
105 100
120 49.2 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 22.978 ug/l
RT: 12.379 min Scan# 1732
Delta R.T. -0.001 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01



Tgt Ion: 75 Resp: 919
Ion Ratio Lower Upper
75 100
53 87.8 85.3 127.9
89 43.5 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#82
4-Chlorotoluene
Concen: 20.658 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion: 91 Resp: 78616
Ion Ratio Lower Upper
91 100
126 37.0 17.0 51.0









#83
tert-Butylbenzene
Concen: 20.975 ug/l
RT: 13.074 min Scan# 1846
Delta R.T. -0.001 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument : MSVOA_Y
ClientSampleId : VY0421SBS01

Tgt Ion:119 Resp: 82218
Ion Ratio Lower Upper
119 100
91 63.3 34.8 104.4
134 26.7 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#84
1,2,4-Trimethylbenzene
Concen: 21.036 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion:105 Resp: 94243
Ion Ratio Lower Upper
105 100
120 45.4 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 20.823 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY008443.D
 Acq: 21 Apr 2022 15:59




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0421SBS01




Tgt Ion:105 Resp: 123180
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.2 30.4

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 04/22/2022
 Supervised By :Semsettin Yesilyurt 04/22/2022




#86
 p-Isopropyltoluene
 Concen: 20.761 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY008443.D
 Acq: 21 Apr 2022 15:59



Tgt Ion:119 Resp: 102851
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.2 39.5
 91 22.9 12.4 37.2









#87
1,3-Dichlorobenzene
Concen: 20.704 ug/l
RT: 13.367 min Scan# 1191
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion:146 Resp: 54992
Ion Ratio Lower Upper
146 100
111 39.0 19.5 58.5
148 63.6 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#88
1,4-Dichlorobenzene
Concen: 20.842 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Tgt Ion:146 Resp: 54838
Ion Ratio Lower Upper
146 100
111 39.0 19.3 57.8
148 64.2 32.6 97.8









#89
n-Butylbenzene
Concen: 21.267 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59

Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01

Tgt Ion: 91 Resp: 9562
Ion Ratio Lower Upper
91 100
92 51.7 26.3 78.9
134 26.9 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#90
Hexachloroethane
Concen: 20.904 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59


Tgt Ion:117 Resp: 18619
Ion Ratio Lower Upper
117 100
201 69.4 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 21.151 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:146 Resp: 49954
Ion Ratio Lower Upper
146 100
111 39.5 20.3 60.9
148 63.3 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 22.271 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion: 75 Resp: 4068
Ion Ratio Lower Upper
75 100
155 91.1 38.0 113.9
157 114.8 47.8 143.3





#93
1,2,4-Trichlorobenzene
Concen: 21.624 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:180 Resp: 28428
Ion Ratio Lower Upper
180 100
182 97.5 47.3 141.9
145 29.8 14.6 43.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022




#94
Hexachlorobutadiene
Concen: 21.937 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Tgt Ion:225 Resp: 14071
Ion Ratio Lower Upper
225 100
223 64.2 30.8 92.3
227 63.3 31.6 94.7





#95
Naphthalene
Concen: 22.543 ug/l
RT: 15.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59




Instrument :
MSVOA_Y
ClientSampleId :
VY0421SBS01




Tgt Ion:128 Resp: 67184
Ion Ratio Lower Upper
128 100
127 12.8 11.0 16.4
129 11.0 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/22/2022
Supervised By :Semsettin Yesilyurt 04/22/2022



#96
1,2,3-Trichlorobenzene
Concen: 22.583 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008443.D
Acq: 21 Apr 2022 15:59



Tgt Ion:180 Resp: 25985
Ion Ratio Lower Upper
180 100
182 93.6 47.3 141.9
145 32.1 15.3 45.9

