



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042125\
Data File : VY021934.D
Acq On : 21 Apr 2025 08:38
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Title : SW846 8260
Last Update : Fri Mar 28 02:30:29 2025



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	16342	PASS
75	95	30	60	51.5	43824	PASS
95	95	100	100	100.0	85117	PASS
96	95	5	9	6.6	5614	PASS
173	174	0.00	2	1.8	1333	PASS
174	95	50	100	88.8	75557	PASS
175	174	5	9	7.6	5754	PASS
176	174	95	101	96.5	72941	PASS
177	176	5	9	6.7	4921	PASS