

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
 Data File : VY004578.D
 Acq On : 22 Apr 2021 12:32
 Operator : SY/MD
 Sample : M2128-01
 Misc : 6.66G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1

Quant Time: Apr 23 09:18:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	123309	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	219737	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	206795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	83687	61.08	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.16%	
35) Dibromofluoromethane	7.734	113	67820	55.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.24%	
50) Toluene-d8	10.185	98	269028	51.87	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.74%	
62) 4-Bromofluorobenzene	12.483	95	91778	52.14	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
Data File : VY004578.D
Acq On : 22 Apr 2021 12:32
Operator : SY/MD
Sample : M2128-01
Misc : 6.66G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1

Quant Time: Apr 23 09:18:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

