

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
 Data File : VY004581.D
 Acq On : 22 Apr 2021 13:42
 Operator : SY/MD
 Sample : M2107-01
 Misc : 5.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-29-0-2.0

Quant Time: Apr 23 09:19:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	124263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	220825	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	208760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89521	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	81077	58.72	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.44%
35) Dibromofluoromethane	7.734	113	68316	55.75	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.50%
50) Toluene-d8	10.185	98	274560	52.67	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.34%
62) 4-Bromofluorobenzene	12.489	95	88856	50.23	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
Data File : VY004581.D
Acq On : 22 Apr 2021 13:42
Operator : SY/MD
Sample : M2107-01
Misc : 5.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-29-0-2.0

Quant Time: Apr 23 09:19:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

