

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
 Data File : VY004583.D
 Acq On : 22 Apr 2021 14:28
 Operator : SY/MD
 Sample : M2111-03
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-3

Quant Time: Apr 23 09:20:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	120840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	215382	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	201331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	88130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	77939	58.05	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.10%	
35) Dibromofluoromethane	7.734	113	66762	55.86	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.72%	
50) Toluene-d8	10.185	98	266851	52.49	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.98%	
62) 4-Bromofluorobenzene	12.483	95	86304	50.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.04%	
Target Compounds						Qvalue
20) Methylene Chloride	4.735	84	8650	6.67	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
Data File : VY004583.D
Acq On : 22 Apr 2021 14:28
Operator : SY/MD
Sample : M2111-03
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-3

Quant Time: Apr 23 09:20:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

