

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
 Data File : VY004589.D
 Acq On : 22 Apr 2021 16:48
 Operator : SY/MD
 Sample : M2128-02
 Misc : 7.63G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Apr 23 09:21:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	119652	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	212775	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	201571	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92008	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80520	60.57	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.14%	
35) Dibromofluoromethane	7.728	113	66682	56.47	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.94%	
50) Toluene-d8	10.185	98	265908	52.94	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.88%	
62) 4-Bromofluorobenzene	12.483	95	86671	50.85	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.70%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042221\
Data File : VY004589.D
Acq On : 22 Apr 2021 16:48
Operator : SY/MD
Sample : M2128-02
Misc : 7.63G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Apr 23 09:21:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

