

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008454.D
 Acq On : 22 Apr 2022 09:59
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 23 01:49:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	111702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	190246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	179770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89607	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	61533	46.818	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.640%
35) Dibromofluoromethane	7.716	113	58369	47.539	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.080%
50) Toluene-d8	10.179	98	212340	43.683	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.360%
62) 4-Bromofluorobenzene	12.483	95	76526	46.287	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.580%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	48423	47.737	ug/l	99
3) Chloromethane	2.113	50	79882	50.046	ug/l	96
4) Vinyl Chloride	2.253	62	105276	50.759	ug/l	96
5) Bromomethane	2.650	94	90548	46.953	ug/l	97
6) Chloroethane	2.796	64	69778	50.571	ug/l	99
7) Trichlorofluoromethane	3.125	101	103661	48.247	ug/l	91
8) Diethyl Ether	3.528	74	35390	47.858	ug/l	73
9) 1,1,2-Trichlorotrifluo...	3.899	101	63540	49.679	ug/l	94
10) Methyl Iodide	4.095	142	79827	47.460	ug/l	91
11) Tert butyl alcohol	4.954	59	31433	234.648	ug/l	96
12) 1,1-Dichloroethene	3.875	96	60046	49.728	ug/l	81
13) Acrolein	3.729	56	34528	246.750	ug/l	97
14) Allyl chloride	4.479	41	80854	51.301	ug/l #	90
15) Acrylonitrile	5.161	53	87994	232.730	ug/l	98
16) Acetone	3.942	43	74256	242.564	ug/l	93
17) Carbon Disulfide	4.192	76	182487	49.989	ug/l	100
18) Methyl Acetate	4.479	43	38338	44.656	ug/l #	83
19) Methyl tert-butyl Ether	5.216	73	185614	48.581	ug/l	94
20) Methylene Chloride	4.710	84	71913	44.489	ug/l #	80
21) trans-1,2-Dichloroethene	5.216	96	72567	49.267	ug/l	83
22) Diisopropyl ether	6.112	45	186926	49.113	ug/l #	93
23) Vinyl Acetate	6.058	43	650365	249.238	ug/l #	88
24) 1,1-Dichloroethane	6.015	63	117837	49.738	ug/l	97
25) 2-Butanone	6.984	43	113426	239.496	ug/l #	78
26) 2,2-Dichloropropane	6.978	77	112906	51.455	ug/l	94
27) cis-1,2-Dichloroethene	6.984	96	84292	50.030	ug/l	83
28) Bromochloromethane	7.332	49	43685	45.712	ug/l #	75
29) Tetrahydrofuran	7.350	42	72092	236.337	ug/l #	75
30) Chloroform	7.509	83	130140	50.000	ug/l	99
31) Cyclohexane	7.783	56	104946	48.571	ug/l #	85
32) 1,1,1-Trichloroethane	7.704	97	117424	51.523	ug/l	95
36) 1,1-Dichloropropene	7.917	75	99332	50.860	ug/l	97
37) Ethyl Acetate	7.070	43	51213	47.414	ug/l #	90
38) Carbon Tetrachloride	7.899	117	105268	51.876	ug/l	100
39) Methylcyclohexane	9.185	83	131919	52.753	ug/l #	85
40) Benzene	8.161	78	290797	51.036	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008454.D
 Acq On : 22 Apr 2022 09:59
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 23 01:49:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.344	41	67506	128.878	ug/l #	17
42) 1,2-Dichloroethane	8.240	62	85128	49.624	ug/l	91
43) Isopropyl Acetate	8.271	43	98284	49.476	ug/l #	87
44) Trichloroethene	8.941	130	80755	51.443	ug/l	94
45) 1,2-Dichloropropane	9.216	63	70027	51.179	ug/l	95
46) Dibromomethane	9.307	93	47641	49.697	ug/l	93
47) Bromodichloromethane	9.496	83	103871	51.043	ug/l	98
48) Methyl methacrylate	9.295	41	41773	49.457	ug/l #	76
49) 1,4-Dioxane	9.295	88	12696	1050.283	ug/l #	74
51) 4-Methyl-2-Pentanone	10.069	43	256670	245.617	ug/l #	84
52) Toluene	10.246	92	193524	51.358	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	112981	51.772	ug/l	98
54) cis-1,3-Dichloropropene	9.929	75	123678	51.142	ug/l #	83
55) 1,1,2-Trichloroethane	10.642	97	65234	49.858	ug/l	97
56) Ethyl methacrylate	10.508	69	86127	50.982	ug/l #	76
57) 1,3-Dichloropropane	10.788	76	105942	49.603	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.783	63	227980	246.235	ug/l	90
59) 2-Hexanone	10.831	43	182883	254.314	ug/l	83
60) Dibromochloromethane	10.984	129	77385	51.480	ug/l	100
61) 1,2-Dibromoethane	11.087	107	64788	50.065	ug/l	99
64) Tetrachloroethene	10.721	164	69315	51.994	ug/l	97
65) Chlorobenzene	11.514	112	210559	51.801	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	78650	52.408	ug/l	97
67) Ethyl Benzene	11.593	91	372885	52.858	ug/l	98
68) m/p-Xylenes	11.703	106	300415	106.314	ug/l	91
69) o-Xylene	12.032	106	143681	52.816	ug/l	91
70) Styrene	12.044	104	241461	52.669	ug/l	94
71) Bromoform	12.209	173	49069	52.174	ug/l #	98
73) Isopropylbenzene	12.331	105	369068	54.323	ug/l	99
74) N-amyl acetate	12.142	43	89917	53.564	ug/l #	82
75) 1,1,2,2-Tetrachloroethane	12.581	83	78968	52.021	ug/l	100
76) 1,2,3-Trichloropropane	12.636	75	99273	85.605	ug/l #	100
77) Bromobenzene	12.611	156	86558	52.240	ug/l	96
78) n-propylbenzene	12.672	91	447809	54.951	ug/l	98
79) 2-Chlorotoluene	12.758	91	244558	53.321	ug/l	96
80) 1,3,5-Trimethylbenzene	12.813	105	301977	53.549	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.380	75	28062	56.138	ug/l #	82
82) 4-Chlorotoluene	12.855	91	250486	52.702	ug/l	96
83) tert-Butylbenzene	13.075	119	265327	54.199	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	302893	54.133	ug/l	97
85) sec-Butylbenzene	13.251	105	406293	54.993	ug/l	100
86) p-Isopropyltoluene	13.367	119	340424	55.020	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	172840	52.103	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	170425	51.863	ug/l	98
89) n-Butylbenzene	13.696	91	312362	55.623	ug/l	99
90) Hexachloroethane	13.965	117	63235	56.847	ug/l	81
91) 1,2-Dichlorobenzene	13.739	146	150352	50.974	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.355	75	11819	51.809	ug/l	79
93) 1,2,4-Trichlorobenzene	15.007	180	85556	52.109	ug/l	99
94) Hexachlorobutadiene	15.111	225	43048	53.736	ug/l	99
95) Naphthalene	15.239	128	191538	51.460	ug/l	98
96) 1,2,3-Trichlorobenzene	15.428	180	74513	51.850	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
Data File : VY008454.D
Acq On : 22 Apr 2022 09:59
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :


Quant Time: Apr 23 01:49:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 23 01:49:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 111702
Ion Ratio Lower Upper
168 100
99 55.8 46.9 70.3



#2
Dichlorodifluoromethane
Concen: 47.737 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59


Tgt Ion: 85 Resp: 48423
Ion Ratio Lower Upper
85 100
87 31.4 16.1 48.2








#3
Chloromethane
Concen: 50.046 ug/l
RT: 2.113 min Scan# 48
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 50 Resp: 79882
Ion Ratio Lower Upper
50 100
52 32.9 24.6 37.0




#4
Vinyl Chloride
Concen: 50.759 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Tgt Ion: 62 Resp: 105276
Ion Ratio Lower Upper
62 100
64 33.3 25.0 37.6








#5
Bromomethane
Concen: 46.953 ug/l
RT: 2.650 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 94 Resp: 90548
Ion Ratio Lower Upper
94 100
96 94.5 73.5 110.3



#6
Chloroethane
Concen: 50.571 ug/l
RT: 2.796 min Scan# 160
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59


Tgt Ion: 64 Resp: 69778
Ion Ratio Lower Upper
64 100
66 32.1 25.4 38.0








#7
 Trichlorofluoromethane
 Concen: 48.247 ug/l
 RT: 3.125 min Scan# 214
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Instrument :
 MSVOA_Y
 ClientSampleId :







Tgt Ion:101 Resp: 103661
 Ion Ratio Lower Upper
 101 100
 103 66.8 48.0 72.0



#8
 Diethyl Ether
 Concen: 47.858 ug/l
 RT: 3.528 min Scan# 280
 Delta R.T. 0.001 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion: 74 Resp: 35390
 Ion Ratio Lower Upper
 74 100
 45 81.8 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.679 ug/l
RT: 3.899 min Scan# 341
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 63540
Ion Ratio Lower Upper
101 100
85 45.7 40.6 60.8
151 76.6 64.9 97.3



#10
Methyl Iodide
Concen: 47.460 ug/l
RT: 4.095 min Scan# 373
Delta R.T. 0.007 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion:142 Resp: 79827
Ion Ratio Lower Upper
142 100
127 44.4 41.5 62.3
141 14.3 11.9 17.9









#11
Tert butyl alcohol
Concen: 234.648 ug/l
RT: 4.954 min Scan# 513
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 59 Resp: 31433
Ion Ratio Lower Upper
59 100
57 9.4 6.4 9.6



#12
1,1-Dichloroethene
Concen: 49.728 ug/l
RT: 3.875 min Scan# 337
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 96 Resp: 60046
Ion Ratio Lower Upper
96 100
61 146.3 145.8 218.6
98 62.7 52.5 78.7





#13

Acrolein

Concen: 246.750 ug/l

RT: 3.729 min Scan# 313

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 56 Resp: 34528

Ion Ratio Lower Upper

56 100

55 69.6 57.6 86.4



#14

Allyl chloride

Concen: 51.301 ug/l

RT: 4.479 min Scan# 436

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 41 Resp: 80854


Ion Ratio Lower Upper

41 100

39 75.3 58.8 88.2


76 46.6 25.1 37.7#








#15
Acrylonitrile
Concen: 232.730 ug/l
RT: 5.161 min Scan# 548
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion: 53 Resp: 87994




Ion	Ratio	Lower	Upper
53	100		
52	81.3	67.0	100.4
51	36.3	29.8	44.6




#16
Acetone
Concen: 242.564 ug/l
RT: 3.942 min Scan# 348
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 43 Resp: 74256

Ion	Ratio	Lower	Upper
43	100		
58	31.0	22.0	33.0









#17
Carbon Disulfide
Concen: 49.989 ug/l
RT: 4.192 min Scan# 389
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 76 Resp: 182487
Ion Ratio Lower Upper
76 100
78 9.2 7.4 11.0



#18
Methyl Acetate
Concen: 44.656 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.007 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 43 Resp: 38338
Ion Ratio Lower Upper
43 100
74 28.4 16.6 24.8#





#19

Methyl tert-butyl Ether

Concen: 48.581 ug/l

RT: 5.216 min Scan# 51

Delta R.T. -0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 73 Resp: 185614

Ion Ratio Lower Upper

73 100

57 20.2 18.6 27.8



#20

Methylene Chloride

Concen: 44.489 ug/l

RT: 4.710 min Scan# 474

Delta R.T. -0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Tgt Ion: 84 Resp: 71913




Ion Ratio Lower Upper


84 100

49 105.1 114.4 171.6#

51 33.3 35.3 52.9#


86 66.6 53.7 80.5








#21
trans-1,2-Dichloroethene
Concen: 49.267 ug/l
RT: 5.216 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 96 Resp: 72567
Ion Ratio Lower Upper
96 100
61 120.6 120.5 180.7
98 63.7 49.8 74.8



#22
Diisopropyl ether
Concen: 49.113 ug/l
RT: 6.112 min Scan# 704
Delta R.T. -0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 45 Resp: 186926
Ion Ratio Lower Upper
45 100
43 63.3 49.3 73.9
87 33.4 19.2 28.8#
59 12.8 8.4 12.6#









#23
 Vinyl Acetate
 Concen: 249.238 ug/l
 RT: 6.058 min Scan# 695
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 43 Resp: 650365
 Ion Ratio Lower Upper
 43 100
 86 13.4 7.1 10.7#



#24
 1,1-Dichloroethane
 Concen: 49.738 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. -0.006 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion: 63 Resp: 117837
 Ion Ratio Lower Upper
 63 100
 98 8.2 3.6 10.8
 100 5.2 2.0 5.8





#25

2-Butanone

Concen: 239.496 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 43 Resp: 113426

Ion Ratio Lower Upper

43 100

72 32.7 17.7 26.5#



#26

2,2-Dichloropropane

Concen: 51.455 ug/l

RT: 6.978 min Scan# 846

Delta R.T. 0.000 min

Lab File: VY008454.D




Acq: 22 Apr 2022 09:59


Tgt Ion: 77 Resp: 112906

Ion Ratio Lower Upper

77 100

97 23.8 10.4 31.4





#27

cis-1,2-Dichloroethene

Concen: 50.030 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 96 Resp: 84292

Ion Ratio Lower Upper

96 100

61 128.7 0.0 319.8

98 64.6 0.0 127.0



#28

Bromochloromethane

Concen: 45.712 ug/l

RT: 7.332 min Scan# 904

Delta R.T. -0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 49 Resp: 43685


Ion Ratio Lower Upper

49 100

129 2.5 0.0 2.2#

130 89.6 55.4 83.0#





#29

Tetrahydrofuran

Concen: 236.337 ug/l

RT: 7.350 min Scan# 906

Delta R.T. 0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 42 Resp: 72092

Ion Ratio Lower Upper

42 100

72 53.9 30.3 45.5#

71 50.3 29.0 43.4#



#30

Chloroform

Concen: 50.000 ug/l

RT: 7.509 min Scan# 933

Delta R.T. 0.001 min

Lab File: VY008454.D




Acq: 22 Apr 2022 09:59


Tgt Ion: 83 Resp: 130140

Ion Ratio Lower Upper

83 100


85 64.3 51.8 77.6








#31
Cyclohexane
Concen: 48.571 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 56 Resp: 104946
Ion Ratio Lower Upper
56 100
69 34.9 26.7 40.1
84 102.2 67.2 100.8#



#32
1,1,1-Trichloroethane
Concen: 51.523 ug/l
RT: 7.704 min Scan# 965
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 97 Resp: 117424
Ion Ratio Lower Upper
97 100
99 63.9 50.7 76.1
61 40.1 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 46.818 ug/l

RT: 8.143 min Scan# 1037

Delta R.T. 0.001 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 65 Resp: 61533

Ion Ratio Lower Upper

65 100

67 53.5 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 114 Resp: 190246


Ion Ratio Lower Upper

114 100

63 17.8 0.0 43.2


88 14.7 0.0 31.8








#35
 Dibromofluoromethane
 Concen: 47.539 ug/l
 RT: 7.716 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59


Instrument :
 MSVOA_Y
 ClientSampleId :





Tgt Ion:113 Resp: 58369
 Ion Ratio Lower Upper
 113 100
 111 103.1 82.6 124.0
 192 18.4 17.6 26.4




#36
 1,1-Dichloropropene
 Concen: 50.860 ug/l
 RT: 7.917 min Scan# 1000
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59



Tgt Ion: 75 Resp: 99332
 Ion Ratio Lower Upper
 75 100
 110 37.3 17.1 51.2
 77 30.2 24.4 36.6





#37

Ethyl Acetate

Concen: 47.414 ug/l

RT: 7.070 min Scan# 861

Delta R.T. 0.001 min


Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 43 Resp: 51213

Ion Ratio Lower Upper

43 100

61 16.7 9.7 14.5#

70 11.8 7.1 10.7#



#38

Carbon Tetrachloride

Concen: 51.876 ug/l

RT: 7.899 min Scan# 997

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 117 Resp: 105268


Ion Ratio Lower Upper

117 100

119 97.7 77.9 116.9

121 30.7 24.8 37.2









#39
Methylcyclohexane
Concen: 52.753 ug/l
RT: 9.185 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 131919
Ion Ratio Lower Upper
83 100
55 64.7 66.5 99.7#
98 51.7 37.9 56.9



#40
Benzene
Concen: 51.036 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59


Tgt Ion: 78 Resp: 290797
Ion Ratio Lower Upper
78 100
77 23.5 20.4 30.6








#41
Methacrylonitrile
Concen: 128.878 ug/l
RT: 7.344 min Scan# 90
Delta R.T. 0.037 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 41 Resp: 67506
Ion Ratio Lower Upper
41 100
39 25.8 92.9 139.3#
67 0.0 0.0 0.0
52 0.0 0.0 0.0



#42
1,2-Dichloroethane
Concen: 49.624 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 62 Resp: 85128
Ion Ratio Lower Upper
62 100
98 10.5 0.0 14.8









#43
Isopropyl Acetate
Concen: 49.476 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 98284
Ion Ratio Lower Upper
43 100
61 31.4 20.2 30.2#
87 16.8 8.9 13.3#



#44
Trichloroethene
Concen: 51.443 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion:130 Resp: 80755
Ion Ratio Lower Upper
130 100
95 99.5 0.0 188.2





#45

1,2-Dichloropropane

Concen: 51.179 ug/l

RT: 9.216 min Scan# 11

Delta R.T. 0.001 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 63 Resp: 70027

Ion Ratio Lower Upper

63 100

65 31.7 23.0 34.6



#46

Dibromomethane

Concen: 49.697 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 93 Resp: 47641


Ion Ratio Lower Upper

93 100

95 83.0 64.3 96.5

174 94.6 83.4 125.0









#47
Bromodichloromethane
Concen: 51.043 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 103871
Ion Ratio Lower Upper
83 100
85 63.6 51.9 77.9
127 8.4 6.6 10.0



#48
Methyl methacrylate
Concen: 49.457 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 41 Resp: 41773
Ion Ratio Lower Upper
41 100
69 106.8 58.2 87.2#
39 62.4 47.4 71.0









#49
1,4-Dioxane
Concen: 1050.283 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 88 Resp: 12696
Ion Ratio Lower Upper
88 100
43 24.8 32.8 49.2#
58 52.6 59.8 89.6#



#50
Toluene-d8
Concen: 43.683 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 98 Resp: 212340
Ion Ratio Lower Upper
98 100
100 66.3 50.2 75.2









#51
4-Methyl-2-Pentanone
Concen: 245.617 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 256670
Ion Ratio Lower Upper
43 100
58 44.5 28.1 42.1#



#52
Toluene
Concen: 51.358 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 92 Resp: 193524
Ion Ratio Lower Upper
92 100
91 171.5 142.6 213.8









#53
 t-1,3-Dichloropropene
 Concen: 51.772 ug/l
 RT: 10.465 min Scan# 1418
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 75 Resp: 112981
 Ion Ratio Lower Upper
 75 100
 77 30.6 25.4 38.2



#54
 cis-1,3-Dichloropropene
 Concen: 51.142 ug/l
 RT: 9.929 min Scan# 1330
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59


Tgt Ion: 75 Resp: 123678
 Ion Ratio Lower Upper
 75 100
 77 31.5 25.6 38.4
 39 39.8 47.6 71.4#








#55
 1,1,2-Trichloroethane
 Concen: 49.858 ug/l
 RT: 10.642 min Scan# 1447
 Delta R.T. -0.006 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion: 97 Resp: 65234




Ion	Ratio	Lower	Upper
97	100		
83	85.8	69.8	104.8
85	51.2	44.3	66.5
99	59.8	49.8	74.6




#56
 Ethyl methacrylate
 Concen: 50.982 ug/l
 RT: 10.508 min Scan# 1425
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion: 69 Resp: 86127

Ion	Ratio	Lower	Upper
69	100		
41	55.3	63.4	95.0#
39	32.9	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 49.603 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 76 Resp: 105942

Ion Ratio Lower Upper

76 100

78 32.9 25.0 37.6



#58

2-Chloroethyl Vinyl ether

Concen: 246.235 ug/l

RT: 9.783 min Scan# 1306

Delta R.T. 0.001 min

Lab File: VY008454.D




Acq: 22 Apr 2022 09:59


Tgt Ion: 63 Resp: 227980

Ion Ratio Lower Upper

63 100

106 32.2 21.5 32.3









#59
2-Hexanone
Concen: 254.314 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 182883
Ion Ratio Lower Upper
43 100
58 60.5 24.4 73.4



#60
Dibromochloromethane
Concen: 51.480 ug/l
RT: 10.984 min Scan# 1503
Delta R.T. 0.001 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 129 Resp: 77385
Ion Ratio Lower Upper
129 100
127 77.1 38.7 116.1





#61

1,2-Dibromoethane

Concen: 50.065 ug/l

RT: 11.087 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:107 Resp: 64788

Ion Ratio Lower Upper

107 100

109 93.7 75.4 113.2



#62

4-Bromofluorobenzene

Concen: 46.287 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 95 Resp: 76526


Ion Ratio Lower Upper

95 100

174 86.1 0.0 177.0

176 82.6 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 1

Delta R.T. 0.001 min


Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion:117 Resp: 179770

Ion Ratio Lower Upper

117 100

82 51.0 43.0 64.4

119 31.1 25.4 38.0



#64

Tetrachloroethene

Concen: 51.994 ug/l

RT: 10.721 min Scan# 1460

Delta R.T. 0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Tgt Ion:164 Resp: 69315




Ion Ratio Lower Upper


164 100

166 126.7 100.4 150.6

129 94.9 71.4 107.0

131 91.2 70.8 106.2









#65
Chlorobenzene
Concen: 51.801 ug/l
RT: 11.514 min Scan# 111
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:112 Resp: 210559
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 52.408 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion:131 Resp: 78650
Ion Ratio Lower Upper
131 100
133 96.0 47.9 143.6
119 66.7 30.8 92.3





#67

Ethyl Benzene

Concen: 52.858 ug/l

RT: 11.593 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 372885

Ion Ratio Lower Upper

91 100

106 31.7 24.5 36.7



#68

m/p-Xylenes

Concen: 106.314 ug/l

RT: 11.703 min Scan# 1621

Delta R.T. -0.000 min

Lab File: VY008454.D




Acq: 22 Apr 2022 09:59


Tgt Ion:106 Resp: 300415

Ion Ratio Lower Upper

106 100

91 194.7 166.6 249.8









#69
o-Xylene
Concen: 52.816 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:106 Resp: 143681
Ion Ratio Lower Upper
106 100
91 206.6 110.3 331.0



#70
Styrene
Concen: 52.669 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59


Tgt Ion:104 Resp: 241461
Ion Ratio Lower Upper
104 100
78 46.1 42.4 63.6
103 54.2 44.4 66.6








#71
Bromoform
Concen: 52.174 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:173 Resp: 49069
Ion Ratio Lower Upper
173 100
175 47.9 24.6 74.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion:152 Resp: 89607
Ion Ratio Lower Upper
152 100
115 57.4 28.2 84.7
150 175.8 0.0 347.6









#73
Isopropylbenzene
Concen: 54.323 ug/l
RT: 12.331 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 369068
Ion Ratio Lower Upper
105 100
120 27.1 13.4 40.2



#74
N-ethyl acetate
Concen: 53.564 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 43 Resp: 89917
Ion Ratio Lower Upper
43 100
70 50.9 28.7 43.1#
55 25.8 17.3 25.9
61 27.4 17.0 25.6#





#75

1,1,2,2-Tetrachloroethane

Concen: 52.021 ug/l

RT: 12.581 min Scan# 1765

Delta R.T. -0.000 min

Lab File: VY008454.D





Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion:	83	Resp:	78968
Ion	Ratio	Lower	Upper
83	100		
131	10.3	5.6	16.8
85	64.6	32.3	96.8



#76

1,2,3-Trichloropropane

Concen: 85.605 ug/l




RT: 12.636 min Scan# 1774


Delta R.T. 0.007 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Tgt Ion:	75	Resp:	99273
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 52.240 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:156 Resp: 86558
Ion Ratio Lower Upper
156 100
77 182.2 95.5 286.4
158 96.6 48.0 144.0



#78
n-propylbenzene
Concen: 54.951 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 91 Resp: 447809
Ion Ratio Lower Upper
91 100
120 23.3 11.3 33.9





#79

2-Chlorotoluene

Concen: 53.321 ug/l

RT: 12.758 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 244558

Ion Ratio Lower Upper

91 100

126 35.8 16.9 50.6



#80

1,3,5-Trimethylbenzene

Concen: 53.549 ug/l

RT: 12.813 min Scan# 1803

Delta R.T. 0.001 min

Lab File: VY008454.D




Acq: 22 Apr 2022 09:59


Tgt Ion: 105 Resp: 301977

Ion Ratio Lower Upper

105 100

120 50.0 24.4 73.4









#81
 trans-1,4-Dichloro-2-butene
 Concen: 56.138 ug/l
 RT: 12.380 min Scan# 1732
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 75 Resp: 28062
 Ion Ratio Lower Upper
 75 100
 53 80.9 85.3 127.9#
 89 48.7 37.0 55.4



#82
 4-Chlorotoluene
 Concen: 52.702 ug/l
 RT: 12.855 min Scan# 1810
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion: 91 Resp: 250486
 Ion Ratio Lower Upper
 91 100
 126 36.5 17.0 51.0









#83
 tert-Butylbenzene
 Concen: 54.199 ug/l
 RT: 13.075 min Scan# 1846
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:119 Resp: 265327
 Ion Ratio Lower Upper
 119 100
 91 63.6 34.8 104.4
 134 27.4 13.3 39.9



#84
 1,2,4-Trimethylbenzene
 Concen: 54.133 ug/l
 RT: 13.123 min Scan# 1854
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion:105 Resp: 302893
 Ion Ratio Lower Upper
 105 100
 120 46.6 22.3 66.9





#85

sec-Butylbenzene

Concen: 54.993 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 406293

Ion Ratio Lower Upper

105 100

134 20.4 10.2 30.4



#86

p-Isopropyltoluene

Concen: 55.020 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion:119 Resp: 340424


Ion Ratio Lower Upper

119 100

134 26.4 13.2 39.5

91 22.6 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 52.103 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:146 Resp: 172840
 Ion Ratio Lower Upper
 146 100
 111 39.6 19.5 58.5
 148 63.5 31.6 95.0



#88
 1,4-Dichlorobenzene
 Concen: 51.863 ug/l
 RT: 13.447 min Scan# 1907
 Delta R.T. 0.001 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59

Tgt Ion:146 Resp: 170425
 Ion Ratio Lower Upper
 146 100
 111 38.2 19.3 57.8
 148 63.5 32.6 97.8









#89
n-Butylbenzene
Concen: 55.623 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 312362
Ion Ratio Lower Upper
91 100
92 53.0 26.3 78.9
134 26.7 13.0 38.9



#90
Hexachloroethane
Concen: 56.847 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008454.D
Acq: 22 Apr 2022 09:59

Tgt Ion: 117 Resp: 63235
Ion Ratio Lower Upper
117 100
201 68.3 43.0 128.8





#91

1,2-Dichlorobenzene

Concen: 50.974 ug/l

RT: 13.739 min Scan# 1955

Delta R.T. 0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:146 Resp: 150352

Ion Ratio Lower Upper

146 100

111 40.9 20.3 60.9

148 64.6 32.3 96.9



#92

1,2-Dibromo-3-Chloropropane

Concen: 51.809 ug/l

RT: 14.355 min Scan# 2056

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59




Tgt Ion: 75 Resp: 11819


Ion Ratio Lower Upper

75 100

155 95.0 38.0 113.9


157 115.1 47.8 143.3








#93
 1,2,4-Trichlorobenzene
 Concen: 52.109 ug/l
 RT: 15.007 min Scan# 2163
 Delta R.T. 0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59


Instrument :
 MSVOA_Y
 ClientSampleId :





Tgt Ion:180 Resp: 85556
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.3 141.9
 145 30.5 14.6 43.8




#94
 Hexachlorobutadiene
 Concen: 53.736 ug/l
 RT: 15.111 min Scan# 2180
 Delta R.T. -0.000 min
 Lab File: VY008454.D
 Acq: 22 Apr 2022 09:59



Tgt Ion:225 Resp: 43048
 Ion Ratio Lower Upper
 225 100
 223 63.7 30.8 92.3
 227 63.1 31.6 94.7





#95

Naphthalene

Concen: 51.460 ug/l

RT: 15.239 min Scan# 21

Delta R.T. -0.000 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:128 Resp: 191538

Ion Ratio Lower Upper

128 100

127 12.6 11.0 16.4

129 11.3 8.4 12.6



#96

1,2,3-Trichlorobenzene

Concen: 51.850 ug/l

RT: 15.428 min Scan# 2232

Delta R.T. 0.006 min

Lab File: VY008454.D

Acq: 22 Apr 2022 09:59

Tgt Ion:180 Resp: 74513

Ion Ratio Lower Upper

180 100

182 94.5 47.3 141.9

145 32.7 15.3 45.9

