

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008464.D
 Acq On : 22 Apr 2022 15:01
 Operator : KP/MD
 Sample : N2523-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-01-042122

Quant Time: Apr 23 01:52:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	59217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	108582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	106494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	46480	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	34718	49.828	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.660%
35) Dibromofluoromethane	7.716	113	35027	49.984	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.960%
50) Toluene-d8	10.179	98	132155	47.438	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.880%
62) 4-Bromofluorobenzene	12.483	95	44139	46.776	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
Data File : VY008464.D
Acq On : 22 Apr 2022 15:01
Operator : KP/MD
Sample : N2523-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-01-042122

Quant Time: Apr 23 01:52:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

